

Arterial Catheterization Product

For short term use (<30 days)

Rx only.

Indications for Use:

The Arrow® Arterial Catheterization Device permits access to the peripheral arterial circulation or to other small vessels.

Contraindications:

None known.

Clinical Benefits to be Expected:

Permits access to the peripheral arterial circulation or to other small vessels. Facilitates continuous blood pressure measurement. Facilitates blood gas sampling and analyses.

⚠️ General Warnings and Precautions

Warnings:

1. Sterile, Single use: Do not reuse, reprocess or resterilize. Reuse of device creates a potential risk of serious injury and/or infection which may lead to death. Reprocessing of medical devices intended for single use only may result in degraded performance or a loss of functionality.
2. Read all package insert warnings, precautions and instructions prior to use. Failure to do so may result in severe patient injury or death.
3. To minimize the risk of air embolism and blood loss associated with disconnects use only securely tightened Luer-Lock connections.
4. In brachial procedures, collateral flow cannot be guaranteed, and therefore intravascular clotting can result in tissue necrosis.
5. In radial artery procedures, practitioners must ascertain that definite evidence of collateral ulnar flow exists.
6. Accidental infusions of drugs or therapeutics or pressure injection into an arterial system may result in severe patient injury or death.
7. Clinicians must be aware of complications/undesirable side-effects associated with arterial procedures including, but not limited to:

- septicemia
- vessel wall perforation
- thrombosis
- embolization
- hematoma
- arterial spasm
- tissue necrosis
- hemorrhage
- peripheral ischemia and infarction
- peripheral nerve injury
- air embolism
- site infection
- cellulitis
- catheter related blood stream infection (CRBSI)

Precautions:

1. Do not alter the catheter, guidewire, or any other kit/set component during insertion, use or removal.
2. Procedure must be performed by trained personnel well versed in anatomical landmarks, safe technique, and potential complications.
3. Use standard precautions and follow established institutional policies and procedures.

4. Some disinfectants used at catheter insertion site contain solvents which can weaken the catheter material. Alcohol, acetone, and polyethylene glycol can weaken the structure of polyurethane materials. These agents may also weaken the adhesive bond between catheter stabilization device and skin.
 - Do not use acetone or acetone-alcohol on or near the catheter surface.
 - Do not use alcohol to soak catheter surface or allow alcohol to dwell in a catheter lumen to restore catheter patency or as an infection prevention measure.
 - Do not use polyethylene glycol containing ointments at insertion site.
 - Allow insertion site to dry completely prior to applying dressing.
5. Indwelling catheter should be routinely inspected for patency, security of dressing, and possible migration.

Kits/Sets may not contain all accessory components detailed in these instructions for use. Become familiar with instructions for individual component(s) before beginning the procedure.

Insertion Instructions

A Suggested Procedure: Use Sterile Technique.

Prepare for Insertion:

1. Assess for adequate collateral arterial circulation.
 - Use of ultrasound has been shown to increase success with catheter placement.
2. Prep and drape anticipated insertion site per institutional policies and procedures.
3. Administer local anesthetic per institutional policies and procedures.
 - A Protected Needle/Safety Needle should be used in accordance with manufacturer's instructions for use.

SharpsAway® II Locking Disposal Cup (where provided):

The SharpsAway II Locking Disposal Cup is used for disposal of needles (15 Ga. - 30 Ga.).

- Using one-handed technique, firmly push needles into disposal cup holes (refer to Figure 1).




Figure 1

- Once placed into disposal cup, needles will be automatically secured in place so that they cannot be reused.

⚠️ Precaution: Do not attempt to remove needles that have been placed into SharpsAway II Locking Disposal Cup. These needles are secured in place. Damage may occur to needles if they are forced out of disposal cup.

- Where provided, a foam SharpsAway® system may be utilized by pushing needles into foam after use.

⚠️ **Precaution:** Do not re-use needles after they have been placed into the foam SharpsAway system. Particulate matter may adhere to needle tip.

4. Remove guard. Trial advance and retract guidewire through needle using guidewire handle to ensure proper function.

⚠️ **Precaution:** Prior to insertion, ensure guidewire is returned to the original position before insertion or blood flashback may be inhibited.

Insert Catheter:

5. Enlarge puncture site with scalpel if desired.

⚠️ **Warning:** Do not cut guidewire.

- Engage safety and/or locking feature of scalpel (where provided) when not in use to reduce the risk of sharps injury.

6. Perform arterial puncture. Blood flashback in clear hub of introducer needle indicates successful entry into vessel (refer to Figure 2).

⚠️ **Precaution:** If both vessel walls are punctured, subsequent advancement of guidewire could result in inadvertent sub-arterial placement.




Figure 2

7. Stabilize position of introducer needle and carefully advance guidewire into vessel using guidewire handle (refer to Figure 3). When the guidewire is fully retracted, the tip of the guidewire is located at the needle tip.

⚠️ **Precaution:** Do not advance guidewire unless there is free blood flashback in needle hub.




Figure 3

⚠️ **Warning:** To reduce the risk of guidewire damage, do not retract guidewire against edge of needle while in vessel.

⚠️ **Precaution:** If resistance is encountered during guidewire advancement do not force feed, withdraw entire unit and attempt new puncture.

8. Advance entire placement device a maximum of 1 to 2 mm further into vessel to insure that catheter is seated within the vessel.

9. Firmly hold introducer needle hub in position and advance catheter forward, over guidewire into vessel (refer to Figure 4).




Figure 4

Complete Insertion:

10. Hold catheter in place and remove guidewire assembly. Pulsatile blood flow indicates positive arterial placement.

⚠️ **Precaution:** Do not reinsert needle into catheter; doing so may result in patient injury or catheter damage.

Secure Catheter:

11. Attach stopcock, injection cap or connecting tubing to catheter hub. Secure catheter to patient in preferred manner using suture wings, suture groove or wing clip, where provided.

⚠️ **Warning:** Care should be exercised that the catheter is not inadvertently kinked at the hub area when securing catheter to the patient as this may result in catheter damage, breakage and loss of arterial monitoring capabilities.

⚠️ **Warning:** Do not apply tape, staples, or sutures directly to the catheter body to reduce risk of damaging catheter, impeding catheter flow, or adversely affecting monitoring capabilities. Secure only at indicated stabilization locations.

⚠️ **Precaution:** Avoid placement or securement in an area of flexion.

Catheter Stabilization Device (where provided):

A catheter stabilization device should be used in accordance with manufacturer's instructions for use.

12. Document insertion procedure.

Care and Maintenance:

Dressing:

Dress according to institutional policies, procedures, and practice guidelines. Change immediately if the integrity becomes compromised (e.g. dressing becomes damp, soiled, loosened or no longer occlusive).

⚠️ **Precaution:** Minimize catheter manipulation throughout procedure to maintain proper catheter tip position.

Catheter Patency:

Maintain catheter patency according to institutional policies, procedures and practice guidelines. All personnel who care for patients with peripheral intravascular devices must be knowledgeable about effective management to prolong catheter's dwell time and prevent injury.

Catheter Removal Instructions:

Use aseptic technique per institutional policies and procedures.

1. Remove dressing.

⚠️ **Warning:** Do not use scissors to remove dressing to reduce the risk of cutting the catheter.

2. Remove catheter securement device or sutures being careful not to cut catheter.

3. Remove catheter slowly.

⚠️ **Warning:** Do not use excessive force in removing catheter. If resistance is met on removal, stop and follow institutional policies and procedures for difficult to remove catheters.

⚠️ **Warning:** Exposure of arterial circulation to atmospheric pressure may result in entry of air into circulation.

4. Apply pressure at site after catheter is removed per institutional policies and procedures.

5. Cover site with a sterile occlusive dressing.



















6. Document catheter removal procedure including confirmation that entire catheter length has been removed per institutional policies and procedures.

For reference literature concerning patient assessment, clinician education, insertion technique, and potential complications associated with this procedure, consult standard textbooks, medical literature, and Arrow International LLC website: www.teleflex.com

A pdf copy of this IFU is located at www.teleflex.com/IFU

en

Symbol Glossary: Symbols are in compliance with ISO 15223-1. Some symbols may not apply to this product. Refer to product labeling for symbols that apply specifically to this product.

								
Caution	Medical device	Consult instructions for use	Contains a medicinal substance	Do not reuse	Do not resterilize	Sterilized by ethylene oxide	Single sterile barrier system with protective packaging inside	
								
Single sterile barrier system	Keep away from sunlight	Keep dry	Do not use if package is damaged	Not made with natural rubber latex	Catalogue number	Lot number	Use by	Manufacturer
	<i>Arrow, the Arrow logo, SharpsAway, Teleflex and the Teleflex logo are trademarks or registered trademarks of Teleflex Incorporated or its affiliates, in the U.S. and/or other countries. © 2021 Teleflex Incorporated. All rights reserved.</i>							
Date of manufacture								

Arrow International LLC
 Subsidiary of Teleflex Incorporated
 3015 Carrington Mill Blvd., Morrisville, NC 27560 USA
 USA: 1 866 246 6990 | International: +1 919 544 8000

Teleflex®