

BD Insyte™ Autogard™ Catheter

• Suggestions For Use

CLINICIANS MUST BE TRAINED IN THE PRACTICE OF VENIPUNCTURE AND BE AWARE OF INHERENT DANGERS.

Aseptic technique, proper skin preparation and continued protection of the site are essential. Observe Universal Precautions on ALL patients.

- Remove needle cover in a straight outward motion and inspect catheter unit. Rotate catheter 360°.
- Perform venipuncture.
- Observe blood return. **20, 22, 24 Ga.:** blood return will also be visible through the catheter. Lower and advance catheter unit 1/8".
- Holding the flash chamber stationary, advance the catheter off the needle into the vein.
- Before withdrawing needle from catheter, depress button to retract the needle into the clear safety shield. Immediately discard unit in a puncture resistant, leak proof sharps container.

If needle retraction does not occur, depress button again. Dispose of any unshielded needles immediately. Keep needlepoint away from body and fingers at all times.

- Connect tubing or adapter.

NEVER REINSERT THE NEEDLE INTO THE CATHETER.

DO NOT BEND THE NEEDLE WHILE USING THE PRODUCT

- Secure the catheter and dress site per protocol.

REPORT NEEDLESTICKS IMMEDIATELY AND FOLLOW ESTABLISHED PROTOCOL.

- Percutaneous puncture with a contaminated needle may lead to serious illness such as hepatitis, HIV (AIDS) or other infectious diseases.
- Resheathing needles is hazardous.

NOTE: 18 to 22 gauge are suitable for use with any power injectors rated for a maximum of 300 psi.

NON PYROGENIC STERILE: Unless package has been opened or damaged.

DO NOT use scissors at or near insertion site.

CAUTION: Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.



N-04020-100A (3/08)

BD Medical Systems, Inc., 9450 South State Street - Sandy, Utah 84070 - U.S.A.

BD Insyte™ Autogard™ Catheter

• Suggestions For Use

CLINICIANS MUST BE TRAINED IN THE PRACTICE OF VENIPUNCTURE AND BE AWARE OF INHERENT DANGERS.

Aseptic technique, proper skin preparation and continued protection of the site are essential. Observe Universal Precautions on ALL patients.

- Remove needle cover in a straight outward motion and inspect catheter unit. Rotate catheter 360°.
- Perform venipuncture.
- Observe blood return. **20, 22, 24 Ga.:** blood return will also be visible through the catheter. Lower and advance catheter unit 1/8".
- Holding the flash chamber stationary, advance the catheter off the needle into the vein.
- Before withdrawing needle from catheter, depress button to retract the needle into the clear safety shield. Immediately discard unit in a puncture resistant, leak proof sharps container.

If needle retraction does not occur, depress button again. Dispose of any unshielded needles immediately. Keep needlepoint away from body and fingers at all times.

- Connect tubing or adapter.

NEVER REINSERT THE NEEDLE INTO THE CATHETER.

DO NOT BEND THE NEEDLE WHILE USING THE PRODUCT

- Secure the catheter and dress site per protocol.

REPORT NEEDLESTICKS IMMEDIATELY AND FOLLOW ESTABLISHED PROTOCOL.

- Percutaneous puncture with a contaminated needle may lead to serious illness such as hepatitis, HIV (AIDS) or other infectious diseases.
- Resheathing needles is hazardous.

NOTE: 18 to 22 gauge are suitable for use with any power injectors rated for a maximum of 300 psi.

NON PYROGENIC STERILE: Unless package has been opened or damaged.

DO NOT use scissors at or near insertion site.

CAUTION: Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.



N-04020-100A (3/08)

BD Medical Systems, Inc., 9450 South State Street - Sandy, Utah 84070 - U.S.A.

BD Insyte™ Autogard™ Catheter

• Suggestions For Use

CLINICIANS MUST BE TRAINED IN THE PRACTICE OF VENIPUNCTURE AND BE AWARE OF INHERENT DANGERS.

Aseptic technique, proper skin preparation and continued protection of the site are essential. Observe Universal Precautions on ALL patients.

- Remove needle cover in a straight outward motion and inspect catheter unit. Rotate catheter 360°.
- Perform venipuncture.
- Observe blood return. **20, 22, 24 Ga.:** blood return will also be visible through the catheter. Lower and advance catheter unit 1/8".
- Holding the flash chamber stationary, advance the catheter off the needle into the vein.
- Before withdrawing needle from catheter, depress button to retract the needle into the clear safety shield. Immediately discard unit in a puncture resistant, leak proof sharps container.

If needle retraction does not occur, depress button again. Dispose of any unshielded needles immediately. Keep needlepoint away from body and fingers at all times.

- Connect tubing or adapter.

NEVER REINSERT THE NEEDLE INTO THE CATHETER.

DO NOT BEND THE NEEDLE WHILE USING THE PRODUCT

- Secure the catheter and dress site per protocol.

REPORT NEEDLESTICKS IMMEDIATELY AND FOLLOW ESTABLISHED PROTOCOL.

- Percutaneous puncture with a contaminated needle may lead to serious illness such as hepatitis, HIV (AIDS) or other infectious diseases.
- Resheathing needles is hazardous.

NOTE: 18 to 22 gauge are suitable for use with any power injectors rated for a maximum of 300 psi.

NON PYROGENIC STERILE: Unless package has been opened or damaged.

DO NOT use scissors at or near insertion site.

CAUTION: Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.



N-04020-100A (3/08)

BD Medical Systems, Inc., 9450 South State Street - Sandy, Utah 84070 - U.S.A.

BD Insyte™ Autogard™ Catheter

• Suggestions For Use

CLINICIANS MUST BE TRAINED IN THE PRACTICE OF VENIPUNCTURE AND BE AWARE OF INHERENT DANGERS.

Aseptic technique, proper skin preparation and continued protection of the site are essential. Observe Universal Precautions on ALL patients.

- Remove needle cover in a straight outward motion and inspect catheter unit. Rotate catheter 360°.
- Perform venipuncture.
- Observe blood return. **20, 22, 24 Ga.:** blood return will also be visible through the catheter. Lower and advance catheter unit 1/8".
- Holding the flash chamber stationary, advance the catheter off the needle into the vein.
- Before withdrawing needle from catheter, depress button to retract the needle into the clear safety shield. Immediately discard unit in a puncture resistant, leak proof sharps container.

If needle retraction does not occur, depress button again. Dispose of any unshielded needles immediately. Keep needlepoint away from body and fingers at all times.

- Connect tubing or adapter.

NEVER REINSERT THE NEEDLE INTO THE CATHETER.

DO NOT BEND THE NEEDLE WHILE USING THE PRODUCT

- Secure the catheter and dress site per protocol.

REPORT NEEDLESTICKS IMMEDIATELY AND FOLLOW ESTABLISHED PROTOCOL.

- Percutaneous puncture with a contaminated needle may lead to serious illness such as hepatitis, HIV (AIDS) or other infectious diseases.
- Resheathing needles is hazardous.

NOTE: 18 to 22 gauge are suitable for use with any power injectors rated for a maximum of 300 psi.

NON PYROGENIC STERILE: Unless package has been opened or damaged.

DO NOT use scissors at or near insertion site.

CAUTION: Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.



N-04020-100A (3/08)

BD Medical Systems, Inc., 9450 South State Street - Sandy, Utah 84070 - U.S.A.