

ARROW®

Pressure Injectable Arrowg⁺ard Blue Plus[®] Three-Lumen CVC

REF **CDA-45703-XPCN1A**

7 Fr. | **3** Lumen | **20** cm catheter length | **.032** inch dia. spring-wire guide


Ergo[®]Pack[®] Complete

Contents:

- | | | |
|--|---|--|
| <ul style="list-style-type: none">1: Arrowg⁺ard Blue Plus[®]1 Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 20 cm, Pressure Injectable1: Spring-Wire Guide, Nitinol, Marked: .032" (0.81 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow[®] GlideWheel[™] Wire Advancer1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle1: Pressure Transduction Probe1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow[®] Raulerson Spring-Wire Introduction Syringe1: Injection Needle: Eclipse[™]2 22 Ga. x 1-1/2" (3.81 cm) RW and 5 mL Luer-Slip Syringe1: Injection Needle: Eclipse[™]2 25 Ga. x 1" (2.54 cm) RW and 3 mL Luer-Lock Syringe1: Syringe: 10 mL Luer-Lock | <ul style="list-style-type: none">1: Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm1: SecondSite[™] Adjustable Hub: Fastener1: SecondSite[™] Adjustable Hub: Catheter Clamp1: SharpsAway[®] Disposal Cup1: SharpsAway II[™] Locking Disposal Cup1: Maximal Barrier Drape[™] with 4" fenestration1: Towel1: Needle Holder1: Forceps1: Safety Scalpel: #111: Checklist/CLIP Sheet1: Flow Rate Information Card with Injection Log1: Sterile Procedure Sign2: Gauze Pad: 2" x 2" (5 cm x 5 cm)5: Gauze Pad: 4" x 4" (10 cm x 10 cm) | <ul style="list-style-type: none">1: Surgical Apparel: Impervious Gown1: Dressing: Tegaderm[®]3 CHG 3-1/2" x 4-1/2" (8.5 cm x 11.5 cm)1: Transducer Cover: 14 cm x 147 cm1: Surgical Apparel: Mask with Eye Shield1: Surgical Apparel: Bouffant Cap1: Suture: 3-0 Silk with Curved Needle1: HemoHopper[®] Fluid Receptacle |
|--|---|--|

¹Licensed under US Patent Nos. 6,706,024; 6,872,195 and 7,329,412.
²A trademark of Becton, Dickinson and Company.
³A registered trademark of 3M Company.

Contraindications: The Arrowg⁺ard Blue Plus antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine, and/or sulfa drugs (refer to product instructions for references).



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	MAX Pressure Injection Flow Rate‡ (mL/sec)
Distal (16 Ga.)	0.42	2750	10
Medial (18 Ga.)	0.38	1100	5
Proximal (18 Ga.)	0.40	1200	5

* Priming volumes are approximate and are done without injection caps.
† Based on 100 cm head height.
‡ Pressure injection flow rates are determined using media of 11.8 centipoise (cP) viscosity, with 152 cm administration set tubing.



Not made with natural rubber latex



Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F).

A-45703-311A (03/20)