



Two-Lumen CVC Kit

REF **CDC-12702-B1A**

7 Fr. | **2** Lumen | **16** cm catheter length | **.025** inch dia. spring-wire guide

ErgoPack®

Contents:

- 1: Two-Lumen Indwelling Catheter: 7 Fr. x 16 cm Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps
- 1: Spring-Wire Guide, Marked: .025" (0.64 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- 1: Catheter: 20 Ga. x 1-3/4" (4.45 cm) Radiopaque over 22 Ga. TW Introducer Needle
- 1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) RW
- 1: Injection Needle: SafetyGlide™ 25 Ga. x 1" (2.54 cm)
- 1: Local Anesthetic Component
- 1: Pressure Transduction Probe
- 1: Introducer Needle: Echogenic 20 Ga. x 1-1/2" (3.81 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- 1: Injection Needle: SafetyGlide™ 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe

- 1: Syringe: 3 mL Luer-Lock
- 1: Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm
- 1: 5 mL 1% Lidocaine HCl Solution
- 1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep® One-Step Solution with Hi-Lite Orange™ Tint (non-sterile solution)
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SharpsAway® Disposal Cup
- 1: SharpsAway II™ Locking Disposal Cup
- 1: Maximal Barrier Drape™ with 4" fenestration
- 1: Needle Holder
- 1: Safety Scalpel: #11
- 1: Checklist/CLIP Sheet
- 1: Sterile Procedure Sign

- 1: Medication Label: 1% Lidocaine
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Surgical Apparel: Impervious Gown
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Suture: 3-0 Silk with Curved Needle
- 1: HemoHopper® Fluid Receptacle

¹A trademark of Becton, Dickinson and Company.

²A registered trademark of CareFusion or one of its subsidiaries.

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.



Not made with natural rubber latex



Store between 20-25°C (68-77°F).

