



ARROWgard Blue[®] Two-Lumen Hemodialysis Catheterization Kit for High Volume Infusions



REF **CDC-26142-1A**

14 Fr. | **2** Lumen | **25** cm catheter length | **.035** inch dia. spring-wire guide



Contents:

- 1: Two-Lumen Indwelling Catheter: 14 Fr. x 25 cm Radiopaque Polyurethane with Blue FlexTip[®], ARROWgard[®] Antimicrobial Surface Treatment¹, Extension Line Clamps
- 1: Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- 1: Injection Needle: SafetyGlide^{™2} 25 Ga. x 1" (2.54 cm)
- 1: Local Anesthetic Component
- 1: Pressure Transduction Probe
- 1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow[®] Raulerson Spring-Wire Introduction Syringe
- 1: Injection Needle: SafetyGlide^{™2} 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
- 1: Syringe: 3 mL Luer-Lock

- 1: Tissue Dilator: 14 Fr. (4.7 mm) x 14 cm
- 1: Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm
- 1: 5 mL 1% Lidocaine HCl Solution
- 1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep^{®3} One-Step Solution with Hi-Lite Orange[™] Tint (non-sterile solution)
- 1: SharpsAway[®] Disposal Cup
- 1: SharpsAway II[™] Locking Disposal Cup
- 1: Maximal Barrier Drape[™] with 4" fenestration
- 1: Needle Holder
- 1: Safety Scalpel: #11
- 1: Checklist/CLIP Sheet
- 1: Sterile Procedure Sign
- 1: Medication Label: 1% Lidocaine
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)

- 1: Surgical Apparel: Impervious Gown
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Suture: 3-0 Braided Silk with Straight Needle
- 1: Suture: 3-0 Silk with Curved Needle
- 1: HemoHopper[®] Fluid Receptacle

¹Licensed under US Patent No. 6,706,024.
²A trademark of Becton, Dickinson and Company.
³A registered trademark of CareFusion or one of its subsidiaries.

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Contraindications: The Arrow large-bore two-lumen catheter is not designed for long-term hemodialysis or for use in patients with thrombosed vessels. The ARROWgard Blue antimicrobial surface catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine, and/or sulfa drugs (refer to product instructions for references).


Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.




Not made with natural rubber latex



Store between 20-25°C (68-77°F).



*Priming volumes are approximate and are done without injection caps.



[†] In vitro tests performed using 3:1 glycerine/saline at 37°C.