

REF NA-04220-X1A

ARROW®



# QuickFlash® Radial Artery Catheterization Kit

20<sup>Ga.</sup> | 3.81 cm catheter length | .018 inch dia. spring-wire guide

## Contents:

- 1: QuickFlash® Radial Artery Catheter: 20 Ga. x 1-1/2" (3.81 cm) Radiopaque Polyurethane over 21 Ga. TW Needle with Integral .018" (0.46 mm) dia. Spring-Wire Guide
- 1: Injection Needle: SafetyGlide™ 25 Ga. x 1" (2.54 cm)
- 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Syringe: 3 mL Luer-Lock
- 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep
- 1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep® One-Step Solution with Hi-Lite Orange™ Tint (non-sterile solution)
- 1: SharpsAway® Disposal Cup
- 1: SharpsAway II™ Locking Disposal Cup
- 1: Drape: 36" x 41" (91 cm x 104 cm) with 3" x 5" (8 cm x 13 cm) fenestration
- 1: Towel: 13" x 18" (33 cm x 46 cm)
- 1: Needle Holder
- 1: Safety Scalpel: #11
- 1: Checklist Sheet
- 1: Medication Label: 1% Lidocaine
- 4: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 4: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: Tegaderm® 10 cm x 12 cm
- 1: Dressing: BIOPATCH®
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Suture: 3-0 Silk with Curved Needle

- 1: Tubing: 6" Extension with One-Way Stopcock

<sup>1</sup>A trademark of Becton, Dickinson and Company.

<sup>2</sup>A registered trademark of CareFusion or one of its subsidiaries.

<sup>3</sup>A registered trademark of 3M Company.

<sup>4</sup>A registered trademark of Johnson & Johnson Corporation.

Rx only. **Warning: Read all package insert warnings, cautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.**

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

**California Prop. 65**

**Warning: Cancer and Reproductive Harm** [www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov).

Fluid path components are non-pyrogenic.

**Warning: The results of certain animal experiments have shown phthalates to be potentially toxic to reproduction. Proceeding from the present state of scientific knowledge, risks for male premature infants cannot be excluded in the case of long-term exposure or application. Medical products containing phthalates should be used only temporarily with pregnant women, nursing mothers, babies and infants.**

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.



REF NA-04220-X1A

ARROW®



# QuickFlash® Radial Artery Catheterization Kit

20<sup>Ga.</sup> | 3.81 cm catheter length | .018 inch dia. spring-wire guide

Manufacturer:  
Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

**Teleflex®**  
[www.teleflex.com](http://www.teleflex.com)

IA-04220-140A, Rev. 01 (2020-12)