



# Pressure Injectable PICC Set with Blue FlexTip<sup>®</sup> Catheter and 130 cm Marked Wire

REF **CA-05052-HP**      **5** Fr. | **2** Lumen | **50** cm catheter length | **.018** inch dia. spring-wire guide

**Contents:**

- 1: Two-Lumen Indwelling Catheter: 5 Fr. x 50 cm Radiopaque Polyurethane with Blue FlexTip<sup>®</sup>, T-Port Connector, Extension Line Clamps, Contamination Guard and Placement Wire
- 1: GlideThru<sup>™</sup> Peel-Away Sheath: 5 Fr. x 1-1/4" (3 cm) Radiopaque over 5 Fr. Dilator
- 1: Sheath: 14 Ga. x 2-3/4" (7 cm) Radiopaque Peel-Away over 5 Fr. Dilator
- 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on Both Ends)
- 1: Spring-Wire Guide, Marked: .018" (0.46 mm) dia. x 51-3/16" (130 cm)(Straight Soft Tip on One End - Straight Stiff Tip on Other)
- 1: Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 2: Dust Cap: Non-Vented
- 1: SecondSite<sup>™</sup> Adjustable Hub: Fastener
- 1: SecondSite<sup>™</sup> Adjustable Hub: Catheter Clamp
- 1: SharpsAway II<sup>™</sup> Locking Disposal Cup
- 1: SharpsAway<sup>®</sup> Disposal Cup
- 1: Catheter Trimmer
- 1: Forceps
- 1: Safety Scalpel: #11
- 1: PICC Line Insertion Checklist
- 1: Patient ID Card
- 1: Chart Sticker
- 1: Patient Information Booklet
- 1: Paper Tape Measure
- 1: Tourniquet
- 1: Dressing: Tegaderm<sup>®1</sup> 15.5 cm x 10 cm
- 1: Dressing: STATLOCK<sup>®2</sup> Catheter Stabilization Device
- 1: Tape: Steri-Strip<sup>®1</sup>

<sup>1</sup>A registered trademark of 3M Company.  
<sup>2</sup>A registered trademark of C. R. Bard, Inc.

Rx only. **Warning: Read all package insert warnings, cautions and instructions prior to use. Failure to do so may result in severe patient injury or death.**

**California Prop. 65**


**Warning: Cancer and Reproductive Harm** [www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov).

Fluid path components are non-pyrogenic.

**Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of previous or current thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of patient must be completed to ensure no contraindications exist.**



Not made with natural rubber latex



Lumen	Priming Volume* (mL)	Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.40	540	3106	4
Proximal (18 Ga.)	0.40	600	3490	4

\* Priming volumes are approximate and are done without accessories.  
 † Flow rates were determined using room temperature water, 100 cm head height and represent approximate flow capabilities.  
 ‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.  
 \*\* Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.

Manufacturer:  
 Arrow International, Inc.  
 Subsidiary of Teleflex Incorporated  
 2400 Bernville Road  
 Reading, PA 19605 USA

**Teleflex**<sup>®</sup>  
[www.teleflex.com](http://www.teleflex.com)

IA-05052-1011 (04/19)



# Pressure Injectable PICC Set with Blue FlexTip<sup>®</sup> Catheter and 130 cm Marked Wire

REF **CA-05052-HP**      **5** Fr. | **2** Lumen | **50** cm catheter length | **.018** inch dia. spring-wire guide