



Multi-Lumen CVC Kit




REF ASK-12703-PCA

7 Fr. 3 Lumen 16 cm catheter length 032

Contents:-

- 1: Multi-Lumen Indwelling Catheter: 7 Fr. x 16 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps
- 1: Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- 1: Pressure Transduction Probe
- 1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- 1: Injection Needle: Eclipse^{TM1} 22 Ga. x 1-1/2" (3.81 cm) RW and 5 mL Luer-Slip Syringe
- 1: Injection Needle: Eclipse^{™1} 25 Ga. x 1" (2.54 cm) RW and 3 mL Luer-Lock Syringe
- 1: Syringe: 10 mL Luer-Lock
- 1: Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm
- 1: SecondSite[™] Adjustable Hub: Fastener

- 1: SecondSite[™] Adjustable Hub: Catheter Clamp
- 1: SharpsAway® Disposal Cup
- 1: SharpsAway II[™] Locking Disposal Cup
- 1: Drape: 68" x 110" with 4" fenestration
- 1: Towel
- 1: Needle Holder
- 1: Safety Scalpel: #11
- 1: Flow Rate Information Card with Injection Log
- 1: Checklist/CLIP Sheet
- 1: Sterile Procedure Sign
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: Tegaderm^{®2} CHG 3-1/2" x 4-1/2" (8.5 cm x 11.5 cm)
- 1: Suture: 3-0 Silk with Curved Needle
- 1: HemoHopper® Fluid Receptacle



Catheter Cross Section

Lumen	Priming Volume* (mL)	Flow Rate [†] (mL/hr)	MAX Pressure Injection Flow Rate [‡] (mL/sec)
Distal (16 Ga.)	0.40	3600	10
Medial (18 Ga.)	0.36	1850	5
Proximal (18 Ga.)	0.40	2020	5

Priming volumes are approximate and are done without an injection cap.



IA-12703-106A (11/19)

¹ A trademark of Becton, Dickinson and Company.

² A registered trademark of 3M Company.

Based on 100 cm head height.

Fressure injection flow rates are determined using media of 11.8 centipoise viscosity, with 152 cm administration set tubing.