



ARROW® PICC

REF CDA-35052-HPK1A

ErgoPack®

Contents:

- | | | |
|--|---|---|
| 1: Two-Lumen Indwelling TaperFree™ Catheter:
5 Fr. x 50 cm Pressure Injectable Radiopaque
Polyurethane with Blue FlexTip®, T-Port Connector,
Extension Line Clamps, Contamination Guard and
Placement Wire | 1: Syringe: 10 mL Luer-Lock
1: Syringe: 3 mL Luer-Lock | 1: Chart Sticker
1: Patient Information Booklet |
| 1: GlideThru™ Peel-Away Sheath: 5 Fr. x 2-3/4"
(7 cm) Radiopaque over 5 Fr. Dilator | 2: Dust Caps: Non-Vented | 1: Checklist/CLIP Sheet |
| 1: Spring-Wire Guide, Nitinol, Marked: .018"
(0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip
on One End – Straight Stiff Tip on Other) with
Arrow Advancer™ | 1: SecondSite™ Adjustable Hub: Fastener
1: SecondSite™ Adjustable Hub: Catheter Clamp | 2: Paper Tape Measures |
| 1: Injection Needle: SafetyGlide™ ¹ 25 Ga. x 5/8"
(1.60 cm) RW | 1: SharpsAway II™ Locking Disposal Cup | 1: Tourniquet |
| 1: Safety Introducer Needle: Echogenic 21 Ga. x
2-3/4" (7 cm) TW | 1: Catheter Trimmer | 5: Gauze Pads: 2" x 2" |
| 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4"
(7 cm) TW | 1: Drape: 24" x 36"
1: Drape: 60" x 76"
1: Drape: 36" x 41" with 3" x 5" fenestration | 5: Gauze Pads: 4" x 4" |
| | 2: CSR Wraps | 1: Surgical Apparel: Mask |
| | 1: Towel | 1: Dressing: Tegaderm® ² 15.5 cm x 10 cm |
| | 1: Filter: 5 Micron Straw | 1: Surgical Apparel: Impervious Gown |
| | 1: Safety Scalpel: #11 | 1: Surgical Apparel: Mask with Eye Shield |
| | 1: Flow Rate Information Card with Injection Log | 1: Surgical Apparel: Bouffant Cap |
| | 1: Sterile Procedure Sign | 1: Tape: Steri-Strip® ² |
| | 1: Patient ID Card | 1: HemoHopper® Fluid Receptacle |

¹ A trademark of Becton, Dickinson and Company.² A registered trademark of 3M Company.

All components are CE 0086 unless otherwise noted.

Warning: This product contains a chemical known to the State of California to cause cancer, birth defects or other reproductive harm.

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of patient must be completed to ensure no contraindications exist. See additional labeling for product specific contraindications.

Not made with
natural rubber latex