



# ARROW® PICC powered by ARROW® VPS TipTracker™ Stilet



REF CDA-35541-TTS

4 Fr. | 55 cm catheter length | .018 inch dia. spring-wire guide



**Contents:**

- 1: Indwelling TaperFree™ Catheter: 4 Fr. x 55 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps, Contamination Guard, Arrow® VPS TipTracker™ Stilet and T-Port Connector
- 1: GlideThru™ Peel-Away Sheath: 4 Fr. x 2-3/4" (7 cm) Radiopaque over 4 Fr. Dilator
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer
- 1: Injection Needle: Eclipse™ 25 Ga. x 1" (2.54 cm) RW
- 1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 10 mL Luer-Lock
- 1: Syringe: 3 mL Luer-Lock
- 1: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SharpsAway® Disposal Cup
- 1: SharpsAway II™ Locking Disposal Cup
- 1: Catheter Trimmer
- 1: Towel: 24" x 36" (61 cm x 91 cm)
- 1: Drape: 60" x 76" (152 cm x 193 cm)
- 1: Drape: 36" x 41" (91 cm x 104 cm) with 3" x 5" (8 cm x 13 cm) fenestration
- 1: Towel
- 1: Filter: 5 Micron Straw
- 1: Safety Scalpel: #11

- 1: Patient ID Card
- 1: Patient Information Booklet
- 2: Paper Tape Measure
- 1: Tourniquet
- 1: Foam Electrodes (3 per pack)
- 10: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 10: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: Tegaderm®<sup>2</sup> 15.5 cm x 10 cm
- 1: Surgical Apparel: Impervious Gown
- 1: Dressing: STATLOCK®<sup>3</sup> Catheter Stabilization Device
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask
- 1: Equipment Cover: 13 cm x 244 cm (5" x 96")
- 1: Tape: Steri-Strip®<sup>2</sup>
- 1: HemoHopper® Fluid Receptacle
- 1: T-Piece Cover

<sup>1</sup>A trademark of Becton, Dickinson and Company.  
<sup>2</sup>A registered trademark of 3M Company.  
<sup>3</sup>A registered trademark of C. R. Bard, Inc.


All components are CE 0086 unless otherwise noted.

- nl** ARROW PICC aangedreven door ARROW VPS TipTracker stilet
- fr** CCIP ARROW alimenté par le stilet ARROW VPS TipTracker
- de** ARROW PEZK mit Antrieb durch ARROW VPS TipTracker Mandrin
- it** PICC ARROW con mandrino VPS TipTracker ARROW
- pl** Cewnik PICC ARROW zasilany mandrynem VPS TipTracker ARROW
- pt** PICC ARROW alimentado pelo estilete VPS TipTracker ARROW
- es** Catéter central de inserción periférica ARROW alimentado por el estilete VPS TipTracker de ARROW
- sv** ARROW perfert införd centralkateter drivs av ARROW VPS TipTracker-mandräng

**Contraindications:** The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of previous or current thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist. See additional labeling for product specific contraindications.



Not made with natural rubber latex



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.58	1140	5720	5

\* Priming volumes are approximate and are done without accessories.  
 † Flow rates were determined using room temperature water, 100 cm head height and represent approximate flow capabilities.  
 ‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.  
 \*\* Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.