



PSI Kit for use with 7 - 7.5 Fr. Catheters

Keep dry

Do not use

if package is damaged

REF SB-09803

Contents:

- Sheath: 8.5 Fr. x 4" (10 cm) Arrow-Flex® Radiopaque Polyurethane with Integral 1: Hemostasis Valve/Side Port and Tissue Dilator
- 1:
- Spring-Wire Guide, Marked: .025" (0.64 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow Advancer
- Catheter: 20 Ga. x 1-3/4" (4.45 cm) Radiopaque over 22 Ga. TW Introducer Needle 1:
- Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) 1:
- 1: Injection Needle: 25 Ga. x 1" (2.54 cm) Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) RW 1:
- Pressure Transduction Probe 1:
- Introducer Needle: 20 Ga. x 1-1/2" (3.81 cm) XTW and 5 mL Arrow® Raulerson 1:
- Spring-Wire Introduction Syringe
- 1: Syringe: 3 mL Luer-Lock
- Syringe: 5 mL Luer-Slip Obturator: 8 Fr. 1:
- 1:
- 1: Cath-Gard® Catheter Contamination Shield: 80 cm with Sterile Tape
- 1: SharpsAway® Disposal Cup 1:
- Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration Drape: 24" x 36" (61 cm x 91 cm)
- 1: Scalpel: #11 1:
- 1: Stopcock: 3-way High-Flow
- 2:
- Prep Sponge Swab Gauze Pad: 2" x 2" (5 cm x 5 cm) 2:
- Gauze Pad: 4" x 4" (10 cm x 10 cm) 5:
- 1: Suture: 3-0 Braided Silk with Straight Needle
- 1: Tray: Prep
- i STERILE EO STERINZE Sterilized Consult Do not reuse Do not resterilize Keep away instructions for use by ethylene oxide from sunlight
- Rx only. Warning: Read all package insert warnings, cautions, and instructions prior to use.
- Failure to do so may result in severe patient injury or death. California Prop. 65 Warning: Cancer and Reproductive Harm www.P65Warnings.ca.gov.

This Product Contains Dry Natural Rubber.

Sterilized using ethylene oxide unless otherwise indicated in the contents list. Fluid path components are non-pyrogenic.



Manufacturer: Arrow International, Inc. Subsidiary of Teleflex, Incorporated 2400 Bernville Road Reading, PA 19605 USA





PSI Kit for use with 7 - 7.5 Fr. Catheters



REF SB-09803

8.5^{Fr.} 10^{cm} sheath length

inch dia. spring-wire