

REF ASK-04020-UNC

ARROW®



Radial Artery Catheterization Kit

20^{Ga.} | 4.45 cm catheter length | .018 inch dia. spring-wire guide

Contents:

- 1: Catheter: 20 Ga. x 4.45 cm Radiopaque Polyurethane over 22 Ga. TW Introducer Needle with Integral .018" (0.46 mm) dia. Spring-Wire Guide
- 1: Injection Needle: SafetyGlide™¹ 18 Ga. x 1-1/2" (3.81 cm) RW
- 1: Injection Needle: SafetyGlide™¹ 25 Ga. x 1" (2.54 cm)
- 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Syringe: 3 mL Luer-Lock
- 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep
- 1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®² One-Step Solution with Hi-Lite Orange™ Tint (non-sterile solution)
- 1: Wing Clip: Suture
- 1: SharpsAway II™ Locking Disposal Cup
- 1: Towel: 13" x 18" (33 cm x 46 cm)
- 1: Drape: 21" x 36" (53 cm x 91 cm) with 3" (7.6 cm) fenestration, with adhesive
- 1: Needle Holder
- 1: Safety Scalpel: #11
- 4: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: Tegaderm®³ 7 cm x 6 cm
- 1: Suture: 3-0 Silk with Curved Needle
- 1: Liquid Adhesive: mastisol®⁴
- 1: Tubing: 12" (30 cm) Pressure Monitoring

¹A trademark of Becton, Dickinson and Company.

²A registered trademark of CareFusion or one of its subsidiaries.

³A registered trademark of 3M Company.

⁴A registered trademark of Ferndale Laboratories, Inc.

Rx only. **Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.**

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

⚠ **California Prop. 65**

Warning: Cancer and Reproductive Harm www.P65Warnings.ca.gov.

Fluid path components are non-pyrogenic.



Not made with natural rubber latex



Store between 20 - 25°C (68 - 77°F).

REF ASK-04020-UNC



Radial Artery Catheterization Kit

20^{Ga.} | 4.45 cm catheter length | .018 inch dia. spring-wire guide

ARROW®

Manufacturer:
Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA

Teleflex®
www.teleflex.com

IC-04020-200A, Rev. 01 (2020-12)