

inch dia.

spring-wire quide



Pressure Injectable PICC Kit with Blue FlexTip® Catheter and 130 cm Marked Wire

REF PR-05541-HPX

Contents:

- Indwelling Catheter: 4 Fr. x 55 cm Pressure Injectable Radiopaque Polyurethane 1: with Blue FlexTip® and Extension Line Clamp
- GlideThru[™] Peel-Away Sheath: 4 Fr. x 4" (10 cm) Radiopaque over 4 Fr. Dilator 1:
- Spring-Wire Guide: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on 1: Both Ends)
- Spring-Wire Guide, Marked: .018" (0.46 mm) dia. x 51-3/16" (130 cm)(Straight 1: Soft Tip on One End - Straight Stiff Tip on Other)
- Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW
- Injection Needle: SafetyGlide^{™1} 25 Ga. x 5/8" (1.60 cm) RW 1:
- Injection Needle: SafetyGlide^{™1} 21 Ga. x 1-1/2" (3.81 cm) TW 1:
- Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm) 1:
- Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW 1:
- Syringe: 10 mL Luer-Lock 3:
- 1: Syringe: 3 mL Luer-Lock
- 1: Skin Protectant Prep Pad (non-sterile solution)
- 10 mL Ampule 0.9% Saline Solution 1:
- 1:
- 5 mL 1% Lidocaine HCl Solution and Alcohol Prep (non-sterile solution) 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®2 One-Step Solution with 1: Hi-Lite Orange[™] Tint (non-sterile solution)
- Dust Cap: Non-Vented 1:
- SecondSite[™] Adjustable Hub: Catheter Clamp 1:
- SecondSite[™] Adjustable Hub: Fastener 1:
- SharpsAway II[™] Locking Disposal Cup 1:
- 1: SharpsAway® Disposal Cup

Bx only. Warning: Read all package insert warnings, cautions and instructions prior to use. Failure to do so may result in severe patient injury or death.


Sterilized using ethylene oxide unless otherwise indicated in the contents list.

🕂 California Prop. 65

Warning: Cancer and Reproductive Harm www.P65Warnings.ca.gov.

Fluid path components are non-pyrogenic.





Lumen	Priming Volume* (mL)	Flow Rate [†] (mL/hr)	Pump Flow Rate [‡] (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.50	660	4075	4

STERILE EO

emperature water, 100 cm head height and represent approxim te flow capabilities.

Pump flow rates are determined at maximum pump pressure of 10 psig.

- Catheter Trimmer 1:
- 1: Towel: 24" x 36" (61 cm x 91 cm)
- Particulate Filter: 5 Micron Straw 1:
- Safety Scalpel: #11 1:
- 1: Patient ID Card
- Chart Sticker 1:
- Patient Information Booklet 1:
- 1: Medication Label: 1% Lidocaine
- Medication Label: 0.9% Saline 1:
- Checklist/CLIP Sheet 1:
- 2: Paper Tape Measure
- Tourniquet 1:
- Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5:
- 10: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- Dressing: Tegaderm^{®3} 15.5 cm x 10 cm 1:
- Dressing: BIOPATCH®4 1:
- Dressing: STATLOCK^{®5} Catheter Stabilization Device 1:

4^{Fr.} 55^{cm}_{catheter}

- Tape: Steri-Strip®3 1:
- ¹A trademark of Becton, Dickinson and Company.
- ²A registered trademark of CareFusion or one of its subsidiaries. ³A registered trademark of 3M Company.
- ⁴A registered trademark of Johnson & Johnson Corporation.
- ⁵A registered trademark of C. R. Bard, Inc.

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of previous or current thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of patient must be completed to ensure no contraindications exist.

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.





~ 25°C (77°F)

Not made with natural rubber latex

LÀXEX

Store hetween 20 - 25°C (68 - 77°F).

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



REF PR-05541-HPX

Pressure Injectable PICC Kit with Blue FlexTip[®] Catheter and 130 cm Marked Wire





^{**} Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.