



# ARROW® JACC with Chloragard® Technology Jugular Axillo-subclavian Central Catheter



REF **CDC-43541-JX1A**

**4.5** Fr. | **35** cm catheter length | **.018** inch dia. spring-wire guide

Distal Lumen  
ARROW. VPS<sup>†</sup>  
Compatible



**Contents:**

- 1: Indwelling TaperFree™ CG+™ Antimicrobial/Antithrombogenic Catheter<sup>1</sup>: 4.5 Fr. x 35 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps and Contamination Guard
- 1: GlideThru™ Peel-Away Sheath: 4.5 Fr. x 2-3/4" (7 cm) Radiopaque over 4.5 Fr. Dilator
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer
- 1: Catheter: 20 Ga. x 1-3/4" (4.50 cm) over 22 Ga. Safety Introducer Needle
- 1: Injection Needle: Eclipse™<sup>2</sup> 25 Ga. x 1-1/2" (3.81 cm) RW
- 1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Syringe: 5 mL Luer-Lock
- 1: Syringe: 5 mL Luer-Slip
- 1: Skin Protectant Prep Pad (non-sterile solution)
- 4: 10 mL Luer-Lock Pre-Filled Saline Syringe
- 1: 5 mL 1% Lidocaine HCl Solution
- 2: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®<sup>3</sup> One-Step Solution with HI-Lite Orange™ Tint (non-sterile solution)
- 1: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener

- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SharpsAway® Disposal Cup
- 1: SharpsAway II™ Locking Disposal Cup
- 1: Towel: 24" x 36" (61 cm x 91 cm)
- 1: Maximal Barrier Drape™ with 4" fenestration
- 1: Towel
- 1: Safety Scalpel: #11
- 1: Checklist/CLIP Sheet
- 1: Sterile Procedure Sign
- 1: Patient Information Booklet
- 1: Medication Label: 1% Lidocaine
- 1: Flow Rate Information Card with Injection Log
- 2: Paper Tape Measure
- 10: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 10: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: Tegaderm®<sup>4</sup> 15.5 cm x 10 cm
- 1: Surgical Apparel: Impervious Gown
- 1: Dressing: BIOPATCH®<sup>5</sup>
- 1: Dressing: STATLOCK®<sup>6</sup> CV Plus Catheter Stabilization Device (prep pad solution non-sterile)
- 1: Surgical Apparel: Mask with Eye Shield


- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask
- 1: Transducer Cover: 5" x 48" (13 cm x 122 cm) with Sterile Gel
- 1: HemoHopper® Fluid Receptacle

<sup>1</sup>Licensed under US Patent Nos. 6,706,024; 6,872,195 and 7,329,412.  
<sup>2</sup>A trademark of Becton, Dickinson and Company.  
<sup>3</sup>A registered trademark of CareFusion or one of its subsidiaries.  
<sup>4</sup>A registered trademark of 3M Company.  
<sup>5</sup>A registered trademark of Johnson & Johnson Corporation.  
<sup>6</sup>A registered trademark of C. R. Bard, Inc.

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

**Contraindications: The Arrow Chloragard JACC is contraindicated for patients with known hypersensitivity to chlorhexidine (refer to product instructions for references).**

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.



Lumen	Priming Volume* (mL)	Flow Rate <sup>†</sup> (mL/hr)	Pump Flow Rate <sup>‡</sup> (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (17 Ga.)	0.44	1580	6290	5

\* Priming volumes are approximate and are done without accessories.  
<sup>†</sup> Based on 100 cm head height.  
<sup>‡</sup> Fluid resuscitation flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.  
\*\* Pressure injection flow rates are determined using media of 11.8 centipoise viscosity, with 152 cm administration set tubing.



Not made with natural rubber latex



Store between 20-25°C (68-77°F).