



Caution



Consult instructions for use



Do not reuse



Do not resterilize

STERILE EO

Sterilized by ethylene oxide



Keep away from sunlight



Keep dry



Do not use if package is damaged



Not made with natural rubber latex



Store below 25°C (77°F). Avoid excessive heat above 30°C (86°F).

DuPont™ Tyvek® packaging

Rx only. Warning: Read all package insert warnings, precautions and instructions prior to use. Failure to do so may result in severe patient injury or death.

Warning: This product contains a chemical known to the State of California to cause cancer, birth defects or other reproductive harm.

Contraindications: Clinical assessment of the patient must be completed to ensure no contraindications exist. The Arrow PICC with Chloragard® Technology is contraindicated in the following areas: Patients with known hypersensitivity to chlorhexidine; In presence of device related infections; In presence of previous or current thrombosis in the intended vessel or along the catheterized vessel pathway.

The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist. See additional labeling for product specific contraindications.

Sterile. Single use. Fluid path components are non-pyrogenic.

LOT

Use By

Manufacturer:
Arrow International, Inc.
Subsidiary of Teleflex Incorporated
2400 Bernville Road
Reading, PA 19605 USA

Teleflex®
www.teleflex.com

IC-45552-500C (7/16)



ARROW® PICC with Chloragard® Technology

Peel to OPEN