



Peripherally Inserted Central Catheter Kit with Placement Wire

REF **PK-01351-TW**

3 Fr. | **55** cm catheter length

Contents:

- 1: Indwelling Catheter, Trimmable: 3 Fr. x 55 cm Radiopaque Polyurethane with T-Port Connector, Integral Extension Line, Extension Line Clamp, Integral Suture Wing, Contamination Guard and Placement Wire
- 1: Peel-Away Sheath: 17 Ga. x 1-1/4" (3.18 cm) Radiopaque over 19 Ga. TW Introducer Needle
- 1: Injection Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Syringe: 10 mL Luer-Slip
- 1: Skin Protectant Prep Pad (non-sterile solution)
- 1: 3 mL Applicator 2% CHG and 70% IPA ChloroPrep®¹ One-Step Solution with Hi-Lite Orange™ Tint (non-sterile solution)
- 1: Dust Cap: Non-Vented
- 1: SharpsAway® Disposal Cup
- 1: Catheter Trimmer
- 1: Towel
- 1: Towel: 24" x 36" (61 cm x 91 cm)
- 1: Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration
- 1: Forceps
- 1: Patient ID Card
- 1: Chart Sticker
- 1: Patient Information Booklet
- 1: Dressing: OpSite®² 14 cm x 10 cm
- 2: Paper Tape Measure
- 1: Tourniquet
- 5: Gauze Pad: 2" x 2" (5 cm x 5 cm)

- 6: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: STATLOCK®³ Catheter Stabilization Device
- 2: Surgical Apparel: Mask
- 1: Surgical Tape: Transpore™⁴ 1" x 54" Roll


¹A registered trademark of CareFusion or one of its subsidiaries.
²A registered trademark of T.J. Smith & Nephew, Limited.
³A registered trademark of C. R. Bard, Inc.
⁴A trademark of 3M Company.

Rx only. **Warning: Read all package insert warnings, cautions and instructions prior to use. Failure to do so may result in severe patient injury or death.**

California Prop. 65
Warning: Cancer and Reproductive Harm www.P65Warnings.ca.gov.

Sterilized using ethylene oxide unless otherwise indicated in the contents list.
 Fluid path components are non-pyrogenic.

Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of previous or current thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of patient must be completed to ensure no contraindications exist.



Lumen	Priming Volume* (mL)	Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)
Distal (20 Ga.)	0.30	92	600

* Priming volumes are approximate and are done without accessories.
 † Flow rates are done with normal saline, room temperature, 100 cm head height and represent approximate flow capabilities.
 ‡ Pump flow rates are determined at maximum pump pressure of 10 psig. Above values apply to the untrimmed catheter. Trimming catheter will alter priming volumes and flow rates.

Not made with natural rubber latex
 Store between 15 - 30°C (59 - 86°F). Avoid freezing and excessive heat above 40°C (104°F).

Arrow International, Inc.
 Subsidiary of Teleflex Incorporated
 2400 Bernville Road
 Reading, PA 19605 USA
Teleflex®
www.teleflex.com
 IK-01351-103J (08/19)



Peripherally Inserted Central Catheter Kit with Placement Wire

REF **PK-01351-TW**

3 Fr. | **55** cm catheter length **ARROW®**