



PICC Kit with Blue FlexTip® Catheter and Placement Wire

REF **PK-01451-WSSP**

4 Fr. | **55** cm catheter length | **.018** inch dia. spring-wire guide

Contents:

- 1: Indwelling Catheter: 4 Fr. x 55 cm Radiopaque Polyurethane with Blue FlexTip®, T-Port Connector, Integral Extension Line, Extension Line Clamp, Integral Suture Wing, Contamination Guard and Placement Wire
- 1: GlideThru™ Peel-Away Sheath: 4 Fr. x 2-3/4" (7 cm) Radiopaque over 4 Fr. Dilator
- 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 13-1/8" (33 cm) (Straight Soft Tip on Both Ends)
- 1: Catheter: 22 Ga. x 1-3/4" (4.45 cm) Radiopaque over 25 Ga. RW Introducer Needle
- 1: Introducer Needle: 21 Ga. x 1-1/2" (3.81 cm) TW
- 1: Injection Needle: SafetyGlide™ 21 Ga. x 1-1/2" (3.81 cm) TW
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Injection Needle: SafetyGlide™ 25 Ga. x 5/8" (1.60 cm) and 3 mL Luer-Lock Syringe
- 1: Syringe: 10 mL Luer-Lock
- 1: Syringe: 10 mL Luer-Slip
- 1: Skin Protectant Prep Pad (non-sterile solution)
- 1: 3 mL Applicator 2% CHG and 70% IPA Chloraprep®² One-Step Solution with Hi-Lite Orange™ Tint (non-sterile solution)
- 1: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SharpsAway II™ Locking Disposal Cup
- 1: SharpsAway® Disposal Cup
- 1: Towel

- 1: Towel: 24" x 36" (61 cm x 91 cm)
- 1: Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration
- 1: Forceps
- 1: Safety Scalpel: #11
- 1: Patient ID Card
- 1: Chart Sticker
- 1: Patient Information Booklet
- 1: Dressing: OpSite®³ 14 cm x 10 cm
- 2: Paper Tape Measure
- 1: Tourniquet
- 5: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 6: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: STATLOCK®⁴ Catheter Stabilization Device
- 2: Surgical Apparel: Mask
- 1: Tape: Steri-Strip®⁵

¹A trademark of Becton, Dickinson and Company.
²A registered trademark of CareFusion or one of its subsidiaries.
³A registered trademark of T.J. Smith & Nephew, Limited.
⁴A registered trademark of C. R. Bard, Inc.
⁵A registered trademark of 3M Company.

Rx only. **Warning: Read all package insert warnings, cautions and instructions prior to use. Failure to do so may result in severe patient injury or death.**


Sterilized using ethylene oxide unless otherwise indicated in the contents list.

California Prop. 65

Warning: Cancer and Reproductive Harm www.P65Warnings.ca.gov.

Fluid path components are non-pyrogenic.

Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of previous or current thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of patient must be completed to ensure no contraindications exist.



Lumen	Priming Volume* (mL)	Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)
Distal (18 Ga.)	0.52	850	3700

* Priming volumes are approximate and are done without accessories.
 † Flow rates are done with normal saline, room temperature, 100 cm head height and represent approximate flow capabilities.
 ‡ Pump flow rates are determined at maximum pump pressure of 10 psig.

Arrow International, Inc.
 Subsidiary of Teleflex Incorporated
 2400 Bernville Road
 Reading, PA 19605 USA
Teleflex®
www.teleflex.com

IK-01451-109I (04/19)



PICC Kit with Blue FlexTip® Catheter and Placement Wire

REF **PK-01451-WSSP**

4 Fr. | **55** cm catheter length | **.018** inch dia. spring-wire guide **ARROW®**