



Two-Lumen Hemodialysis Catheterization Kit for High Volume Infusions

REF AK-12122-F

Contents:

- Two-Lumen Indwelling Catheter: 12 Fr. x 16 cm Radiopaque Polyurethane 1: with Blue FlexTip®, Extension Line Clamps
- Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 23-5/8" (60 cm) (Straight 1: Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer 1:
- Needle Injection Needle: 25 Ga. x 1" (2.54 cm) 1:
- 1: Local Anesthetic Component
- Pressure Transduction Probe
- 1:
- Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® 1: Raulerson Spring-Wire Introduction Syringe
- Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe 1:
- Syringe: 3 mL Luer-Lock 1:

- Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm 1:
- 5 mL 1% Lidocaine HCI Solution 1:
- 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®1 One-Step Solution 1: with Hi-Lite Orange[™] Tint (non-sterile solution)

12^{Fr.} 2^{Lumen} 16^{cm}_{catheter} .035^{spring-wire} guide

- 2: Dust Cap: Non-Vented
- SharpsAway® Disposal Cup 1:
- Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration 1:
- 1: Scalpel: #11
- Gauze Pad: 2" x 2" (5 cm x 5 cm) 2:
- Gauze Pad: 4" x 4" (10 cm x 10 cm) 5:
- 1: Suture: 3-0 Braided Silk with Straight Needle

¹A registered trademark of CareFusion or one of its subsidiaries







with thrombosed vessels.





Rx only. Warning: Read all package insert warnings, cautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

California Prop. 65

Warning: Cancer and Reproductive Harm www.P65Warnings.ca.gov.

Fluid path components are non-pyrogenic.



Contraindications: Not designed for long-term hemodialysis or for use in patients



Catheter Cross Section

Lumen	Priming Volume* (mL)	Flow Rate† (mL/min)	Venous Pressure Setting (mmHg)
Distal (12 Ga.)	1.40	220	100
		370	200
Proximal (12 Ga.)	1.30	220	100
		370	200

† Blood flow rates are approximate and were determined by varying the venous pressure setting from 100 to 200 mmHg in an in vitro experiment using blood analog solution.



Manufacturer: Arrow International, Inc. Subsidiary of Teleflex Incorporated 2400 Bernville Road Reading, PA 19605 USA



IK-12122-101H (9/17)

