



Two-Lumen Hemodialysis Catheterization Kit for High Volume Infusions

Lumen

6 cm catheter length of the catheter length o

Contents:

- 1: Two-Lumen Indwelling Catheter: 12 Fr. x 16 cm Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps
- Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One
- End "L" Tip on Other) with Arrow Advancer Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- Injection Needle: 25 Ga. x 1" (2.54 cm)
- Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- Pressure Transduction Probe
- Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
- Syringe: 3 mL Luer-Lock
- Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm
- 5 mL 1% Lidocaine HCl Solution and Alcohol Prep
- 3 mL Applicator 2% CHG and 70% IPA ChloraPrep on One-Step Solution with Hi-Lite Orange™
- Tint (non-sterile solution) Dust Cap: Non-Vented
- SharpsAway® Disposal Cup
- Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration
- Scalpel: #11
- Gauze Pad: 2" x 2" (5 cm x 5 cm)
- Gauze Pad: 4" x 4" (10 cm x 10 cm) Suture: 3-0 Braided Silk with Straight Needle

¹A registered trademark of CareFusion or one of its subsidiaries.


Rx only. Warning: Read all package insert warnings, cautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

⚠ California Prop. 65

Warning: Cancer and Reproductive Harm www.P65Warnings.ca.gov.

Sterilized using ethylene oxide unless otherwise indicated in the contents list. Fluid path components are non-pyrogenic.

Contraindications: The Arrow large-bore two-lumen catheter is not designed for long-term hemodialysis or for use in patients with thrombosed vessels.



Lumen	Priming Volume* (mL)	Flow Rate† (mL/min)	Venous Pressure Setting (mmHg)
Distal (12 Ga.)	1.40	220	100
		370	200
Proximal (12 Ga.)	1.30	220	100
		370	200

* Priming volumes are approximate and are done without accessories. † Blood flow rates are approximate and were determined by varying the venous pressure setting from 100 to 200 mmHg in an in vitro experiment using blood analog solution.





















Manufacturer: Arrow International, Inc. Subsidiary of Teleflex Incorporated 2400 Bernville Road Reading, PA 19605 USA



IK-12122-110A (2020-03)







