



Arrowg+ard Blue[®] Two-Lumen Hemodialysis **Catheterization Kit for High Volume Infusions**



Contents:

- 1: Two-Lumen Indwelling Catheter: 12 Fr. x 20 cm Radiopaque Polyurethane with Blue FlexTip®, Arrowg+ard® Antimicrobial Surface Treatment¹, Extension Line Clamps
- Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 26-3/4" (68 cm) (Straight Soft Tip on One 1:
- End "J" Tip on Other) with Arrow Advancer Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle 1:
- 1: Injection Needle: 25 Ga. x 1" (2.54 cm)
- 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Pressure Transduction Probe
- Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire 1:
- Introduction Syringe Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe 1:
- Syringe: 3 mL Luer-Lock 1:
- 1: Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm
- 5 mL 1% Lidocaine HCI Solution and Alcohol Prep 1:
- 3 mL Applicator 2% CHG and 70% IPA ChloraPrep^{®2} One-Step Solution with Hi-Lite Orange™ 1: Tint (non-sterile solution)
- 2: Dust Cap: Non-Vented
- SharpsAway® Disposal Cup 1:
- 1: Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration
- 1: Scalpel: #11
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- Suture: 3-0 Braided Silk with Straight Needle 1:

¹Licensed under US Patent No. 6.706.024.

²A registered trademark of CareFusion or one of its subsidiaries.

Rx only. Warning: Read all package insert warnings, cautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

A California Prop. 65


Warning: Cancer and Reproductive Harm www.P65Warnings.ca.gov.

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic.

i

Contraindications: The Arrow large-bore two-lumen catheter is not designed for long-term hemodialysis or for use in patients with thrombosed vessels. The Arrowg+ard Blue antimicrobial surface catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine, and/or sulfa drugs (refer to product instructions for references).



STERILE EO

TERINZE

Lumen	Priming Volume* (mL)	Flow Rate† (mL/min)	Venous Pressure Setting (mmHg)
Distal (12 Ga.)	1.60	200	100
		340	200
Proximal (12 Ga.)	1.50	200	100
		340	200

ARR()//

* Priming volumes are approximate and are done without accessories. † Blood flow rates are approximate and were determined by varying the venous pressure setting from 100 to 200 mmHg in an in vitro experiment using blood analog solution.



¦







Not made with natural rubber latex



Manufacturer: Arrow International, Inc. Subsidiary of Teleflex Incorporated 2400 Bernville Road Reading, PA 19605 USA



IK-25122-105A (2020-03)







Arrowg+ard Blue® Two-Lumen Hemodialysis Catheterization Kit for High Volume Infusions

12^{Fr.} 2^{Lumen} 20^{cm} catheter .035^{inch dia.} spring-wire guide