



ARROWg⁺ard Blue[®] Two-Lumen Hemodialysis Catheterization Kit for High Volume Infusions

REF AK-25142-F

Contents: _

- Two-Lumen Indwelling Catheter: 14 Fr. x 20 cm Radiopaque Polyurethane with Blue FlexTip[®], ARROWg¹ard[®] Antimicrobial Surface Treatment¹, Extension Line Clamps
- 1: Spring-Wire Guide, Marked: .038" (0.97 mm) dia. x 27-1/2" (70 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- 1: Injection Needle: 25 Ga. x 1" (2.54 cm)
- 1: Local Anesthetic Component
- 1: Pressure Transduction Probe
- 1: Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
- 1: Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- 1: Syringe: 3 mL Luer-Lock
- 1: Tissue Dilator: 14 Fr. (4.7 mm) x 14 cm
- 1: Tissue Dilator: 16 Fr. (5.3 mm) x 11.4 cm

- 1: 5 mL 1% Lidocaine HCl Solution
- 1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep^{®2} One-Step Solution with Hi-Lite Orange[™] Tint (non-sterile solution)

14^{Fr.} 2^{Lumen} 20^{cm} catheter length .038 spring-wire guide

- 2: Dust Cap: Non-Vented
- 1: SharpsAway[®] Disposal Cup
- 1: Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration
- 1: Needle Holder
- 1: Scalpel: #11
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Suture: 3-0 Silk with Curved Needle
- ¹Licensed under US Patent No. 6,706,024.

²A registered trademark of CareFusion or one of its subsidiaries.









Catheter Cross Section





 Rx only.
 Warning: Read all package insert warnings, cautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.
 Comparison of the severe patient injury or death.

Sterilized using ethylene oxide unless otherwise indicated in the contents list.


A California Prop. 65

Warning: Cancer and Reproductive Harm www.P65Warnings.ca.gov.

Fluid path components are non-pyrogenic.

Contraindications: The Arrow large-bore two-lumen catheter is not designed for long-term hemodialysis or for use in patients with thrombosed vessels.

The ARROWg² ard Blue antimicrobial surface catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine, and/or sulfa drugs (refer to product instructions for references).



* Priming volumes are approximate and are done without accessories. + Venous Pressure settings are approximate and were determined by varying the blood flow rates from 250 to 400 mL/min in an in vitro experiment using blood analog solution.

DuPont" Tyvek[®]

Manufacturer: Arrow International, Inc. Subsidiary of Teleflex Incorporated 2400 Bernville Road Reading, PA 19605 USA

> Teleflex® www.teleflex.com

> > IK-25142-100H (12/17)

