

ARROW®



ARROW® PICC Set

REF PR-34052-HPSL

Contents:

- | | |
|---|---|
| 1: Two-Lumen Indwelling TaperFree™ Catheter: 5 Fr. x 40 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip® and Extension Line Clamps | 1: SecondSite™ Adjustable Hub: Fastener |
| 1: GlideThru™ Peel-Away Sheath: 5 Fr. x 4" (10 cm) Radiopaque over 5 Fr. Dilator | 1: SecondSite™ Adjustable Hub: Catheter Clamp |
| 1: Spring-Wire Guide: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on Both Ends) | 1: Protection Sleeve |
| 1: Spring-Wire Guide, Marked: .018" (0.46 mm) dia. x 51-3/16" (130 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) | 1: Flow Rate Information Card with Injection Log |
| 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW | 1: Patient ID Card |
| 1: Syringe: 5 mL Luer-Lock | 1: Chart Sticker |
| 1: Skin Protectant Prep Pad | 1: Patient Information Booklet |
| 2: Dust Caps: Non-Vented | 1: Checklist/CLIP Sheet |
| | 1: Paper Tape Measure |
| | 1: Dressing: STATLOCK® ¹ Catheter Stabilization Device |

¹A registered trademark of C. R. Bard, Inc.
All components are CE 0086 unless otherwise noted.



Caution



Consult instructions for use



Do not reuse



Do not resterilize



Sterilized by ethylene oxide



Keep away from sunlight



Keep dry



Do not use if package is damaged



Not made with natural rubber latex

Rx only. Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

Warning: This product contains a chemical known to the State of California to cause cancer, birth defects or other reproductive harm.

Sterile. Fluid path components are non-pyrogenic.

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of patient must be completed to ensure no contraindications exist. See additional labeling for product specific contraindications.

EC REP Teleflex Medical
CE 0086 IDA Business and Technology Park
Dublin Road, Athlone
Co. Westmeath, Ireland

Manufacturer:
Arrow International, Inc.
Subsidiary of Teleflex Incorporated
2400 Bernville Road
Reading, PA 19605 USA

Teleflex®
www.teleflex.com

DuPont™ Tyvek®
package

IR-34052-104A, Rev. 2 (9/16)



ARROW® PICC Set

REF PR-34052-HPSL

5 Fr.

2 Lumen

40 cm catheter length

.018 inch dia. spring-wire guide

130 cm Stainless Steel Wire

ARROW®