



# ARROW® PICC Set

## REF PR-34552-HPHNL

### Contents:

- |   |   |
|---|---|
| <ul style="list-style-type: none"> <li>1: Two-Lumen Indwelling TaperFree™ Catheter: 5 Fr. x 45 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps, and Contamination Guard</li> <li>1: GlideThru™ Peel-Away Sheath: 5 Fr. x 4" (10 cm) Radiopaque over 5 Fr. Dilator</li> <li>1: Guidewire, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) Nitinol (Straight Soft Tip on One End - Straight Stiff Tip on Other)</li> <li>1: Guidewire, Marked: .018" (0.46 mm) dia. x 130 cm Hydrophilic-Coated Nitinol with Soft Tip Tungsten Coil (CE 0120)</li> <li>1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW</li> <li>1: Syringe: 10 mL Luer-Lock</li> <li>1: Skin Protectant Prep Pad</li> <li>2: Dust Caps: Non-Vented</li> </ul> | <ul style="list-style-type: none"> <li>1: SecondSite™ Adjustable Hub: Fastener</li> <li>1: SecondSite™ Adjustable Hub: Catheter Clamp</li> <li>1: Catheter Trimmer</li> <li>1: Safety Scalpel: #11</li> <li>1: Flow Rate Information Card with Injection Log</li> <li>1: Patient ID Card</li> <li>1: Chart Sticker</li> <li>1: Patient Information Booklet</li> <li>1: Checklist/CLIP Sheet</li> <li>1: Paper Tape Measure</li> <li>1: Dressing: STATLOCK®<sup>1</sup> Catheter Stabilization Device</li> </ul> |
|---|---|

<sup>1</sup>A registered trademark of C. R. Bard, Inc.  
All components are CE 0086 unless otherwise noted.



Rx only. Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

Warning: This product contains a chemical known to the State of California to cause cancer, birth defects or other reproductive harm.

Sterile. Fluid path components are non-pyrogenic.

**Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of patient must be completed to ensure no contraindications exist. See additional labeling for product specific contraindications.**

**EC REP** Teleflex Medical  
IDA Business and Technology Park  
Dublin Road, Athlone  
Co. Westmeath, Ireland  
**CE**  
0086

**Manufacturer:**  
Arrow International, Inc.  
Subsidiary of Teleflex Incorporated  
2400 Bernville Road  
Reading, PA 19605 USA

**Teleflex®**  
www.teleflex.com