

17 Ga

Catheter Cross Section

- 1: Foil-Laminated Pouch 79% Skin Protectant Prep Pad
- 1: Dust Cap: Non-Vented
- ⑦ 1: Clamp: Catheter
- ⑧ 1: Fastener: Catheter Clamp
 - 1: Chart Sticker
 - 1: Checklist/CLIP Sheet
 - 1: Flow Rate Information Card with Injection Log
 - 1: Patient ID Card
 - 1: Patient Information Booklet
- ③ 1: Dressing: STATLOCK^{®1} Catheter Stabilization Device
 - 1: Paper Tape Measure
- ¹A registered trademark of C. R. Bard, Inc.



Rx only. Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

Sterile. Single use – do not resterilize. Sterilized by ethylene oxide. Do not use if package has been previously opened or damaged. Does not contain natural rubber latex. Fluid path components are non-pyrogenic. Contains no medication.

cm catheter

length

⁺⁺Meets VasoNovaTM VPSTM stylet requirements for use.



2400 Bernville Road Reading, PA 19605 USA www.arrowintl.com

Arrow**ADVANTAGE⁵™ Pressure Injectable** PICC Set

inch dia.

spring-wire guide

MAX Pressure Priming **Pump Flow** Injection Flow Rate** Flow Volume Rate Rate[‡] Lumen (mL)(mL/hr) (mL/hr) (mL/sec) Distal (17 Ga.) 0.55 1220 5740 5


* Priming volumes are approximate and are done without injection caps.

Distal Lumen VasoNova™ VPS compatible#

* Flow rates were determined using room temperature water, 100 cm head height and represent approximate flow capabilities.

* Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

**Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.



IR-35041-101C (6/16)

INTERNATIONAL

REF PR-35041-HPSL