



ARROW® PICC powered by ARROW® VPS TipTracker™ Stylet



REF **PR-35541-TTS**

4 Fr. | **55** cm catheter length | **.018** inch dia. spring-wire guide

Contents:

- 1: Indwelling TaperFree™ Catheter: 4 Fr. x 55 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps, Contamination Guard, Arrow® VPS TipTracker™ Stylet and T-Port Connector
- 1: GlideThru™ Peel-Away Sheath: 4 Fr. x 2-3/4" (7 cm) Radiopaque over 4 Fr. Dilator
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer
- 1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 10 mL Luer-Lock
- 1: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SharpsAway® Disposal Cup
- 1: SharpsAway II™ Locking Disposal Cup
- 1: Catheter Trimmer
- 1: Safety Scalpel: #11
- 1: Patient Information Booklet
- 2: Paper Tape Measure
- 1: Tourniquet
- 1: Foam Electrodes (3 per pack)
- 1: Dressing: STATLOCK®¹ Catheter Stabilization Device
- 1: Equipment Cover: 13 cm x 244 cm (5" x 96")
- 1: T-Piece Cover

¹A registered trademark of C. R. Bard, Inc.


All components are CE 0086 unless otherwise noted.

- nl** ARROW PICC aangedreven door ARROW VPS TipTracker stilet
- fr** CCIP ARROW alimenté par le stylet ARROW VPS TipTracker
- de** ARROW PEZK mit Antrieb durch ARROW VPS TipTracker Mandrin
- it** PICC ARROW con mandrino VPS TipTracker ARROW
- pl** Cewnik PICC ARROW zasilany mandrynem VPS TipTracker ARROW
- pt** PICC ARROW alimentado pelo estilete VPS TipTracker ARROW
- es** Catéter central de inserción periférica ARROW alimentado por el estilete VPS TipTracker de ARROW
- sv** ARROW perifert införd centralkateter drivs av ARROW VPS TipTracker-mandräng

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of previous or current thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of the patient must be completed to ensure no contraindications exist. See additional labeling for product specific contraindications.



Not made with natural rubber latex



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.58	1140	5720	5

* Priming volumes are approximate and are done without accessories.
 † Flow rates were determined using room temperature water, 100 cm head height and represent approximate flow capabilities.
 ‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.
 ** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.