



ARROW® PICC Set

REF **PR-35552-HPHNM**

Contents:

- | | |
|---|---|
| 1: Two-Lumen Indwelling TaperFree™ Catheter: 5 Fr. x 55 cm
Pressure Injectable Radiopaque Polyurethane with Blue FlexTip®,
Extension Line Clamps, and Contamination Guard | 1: SecondSite™ Adjustable Hub: Fastener |
| 1: GlideThru™ Peel-Away Sheath: 5 Fr. x 4" (10 cm)
Radiopaque over 5 Fr. Dilator | 1: SecondSite™ Adjustable Hub: Catheter Clamp |
| 1: Guidewire, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm)
Nitinol (Straight Soft Tip on One End - Straight Stiff Tip on Other) | 1: Catheter Trimmer |
| 1: Guidewire, Marked: .018" (0.46 mm) dia. x 80 cm
Hydrophilic-Coated Nitinol with Soft Tip Tungsten Coil (CE 0120) | 1: Safety Scalpel: #11 |
| 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW | 1: Chart Sticker |
| 1: Syringe: 10 mL Luer-Lock | 1: Checklist/CLIP Sheet |
| 1: Skin Protectant Prep Pad | 1: Flow Rate Information Card with Injection Log |
| 2: Dust Caps: Non-Vented | 1: Patient ID Card |
| | 1: Patient Information Booklet |
| | 1: Dressing: STATLOCK® ¹ Catheter Stabilization Device |
| | 1: Paper Tape Measure |

¹A registered trademark of C. R. Bard, Inc.
All components are CE 0086 unless otherwise noted.



Caution



Consult
instructions for use



Do not reuse



Do not resterilize



Sterilized
by ethylene oxide



Keep away
from sunlight



Keep dry



Do not use
if package is damaged



Not made with
natural rubber latex

Rx only. Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

Warning: This product contains a chemical known to the State of California to cause cancer, birth defects or other reproductive harm.

Sterile. Fluid path components are non-pyrogenic.

Contraindications: The Pressure Injectable PICC is contraindicated wherever there is presence of device related infections or presence of thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of patient must be completed to ensure no contraindications exist. See additional labeling for product specific contraindications.

EC REP Teleflex Medical
CE IDA Business and Technology Park
0086 Dublin Road, Athlone
Co. Westmeath, Ireland

Manufacturer:
Arrow International, Inc.
Subsidiary of Teleflex Incorporated
2400 Bernville Road
Reading, PA 19605 USA

Teleflex®
www.teleflex.com