

ARROW PICC with Chlorag⁺ard[®] Technology

Dust Caps: Non-Vented

Catheter Trimmer

Safety Scalpel: #11

Checklist/CLIP Sheet

Paper Tape Measure

² A registered trademark of C. R. Bard, Inc.

Patient Information Booklet

Chart Sticker

Patient ID Card

REF PR-44563-HPHNL

Contents:

- 1: Multi-Lumen Indwelling TaperFree™ CG^{+™} Antimicrobial/Antithrombogenic Catheter1: 6 Fr. x 45 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip[®], Extension Line Clamps and Contamination Guard
- 1: GlideThru[™] Peel-Away Sheath: 6 Fr. x 4" (10 cm) Radiopaque over 6 Fr. Dilator
- 1: Guidewire, Marked: .018" (0.46 mm) dia. x 130 cm Hydrophilic-Coated Nitinol with Soft Tip Tungsten Coil
- Guidewire, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) 1: Nitinol (Straight Soft Tip on One End – Straight Stiff Tip on Other)
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 10 mL Luer-Lock
- Skin Protectant Prep Pad 1:







STERILE EO



Keep away

3:

1:

1:

1:

1:

1:

1.

1:

1:

1:

1:

Keep dry



¹ Licensed under US Patent Nos. 6,706,024; 6,872,195 and 7,329,412.

Flow Rate Information Card with Injection Log

Dressing: STATLOCK®2 Catheter Stabilization Device





Consult

Sterilized by ethylene oxide

from sunlight

Do not use if package is damaged

Not made with natural rubber latex

SecondSite[™] Adjustable Hub: Catheter Clamp and Fastener

Store below 25°C (77°F). Avoid excessive heat above 30°C (86°F).

instructions for use

Do not reuse



Rx only. Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. Sterile. Fluid path components are non-pyrogenic.

Contraindications: The Arrow® PICC with Chloragtard® Technology is contraindicated for patients with known hypersensitivity to chlorhexidine (refer to product instructions for references).

> Arrow International, Inc. Subsidiary of Teleflex, Incorporated 2400 Bernville Road Reading, PA 19605 USA

> > etiex® www.teleflex.com

> > > IR-44563-100B (5/16)