







ARROW PICC with Chloragtard® Technology

REF PR-45563-HPHNL

Contents:

- 1: Multi-Lumen Indwelling TaperFree™ CG+™ Antimicrobial/Antithrombogenic Catheter1: 6 Fr. x 55 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps and Contamination Guard
- 1: GlideThru™ Peel-Away Sheath: 6 Fr. x 4" (10 cm) Radiopaque over 6 Fr. Dilator
- 1: Guidewire, Marked: .018" (0.46 mm) dia. x 130 cm Hydrophilic-Coated Nitinol with Soft Tip Tungsten Coil
- Guidewire, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) Nitinol (Straight Soft Tip on One End – Straight Stiff Tip on Other)
- Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- Syringe: 10 mL Luer-Lock
- Skin Protectant Prep Pad

- Dust Caps: Non-Vented
- SecondSite[™] Adjustable Hub: Catheter Clamp and Fastener
- Catheter Trimmer
- Safety Scalpel: #11
- Chart Sticker 1:
- Checklist/CLIP Sheet
- Flow Rate Information Card with Injection Log
- Patient ID Card
- Patient Information Booklet
- Dressing: STATLOCK®2 Catheter Stabilization Device
- Paper Tape Measure











by ethylene oxide









if package is damaged



Not made with natural rubber latex



Rx only. Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. Sterile. Fluid path components are non-pyrogenic.

Contraindications: The Arrow® PICC with Chloragtard® Technology is contraindicated for patients with known hypersensitivity to chlorhexidine (refer to product instructions for references).

> Arrow International, Inc. Subsidiary of Teleflex, Incorporated 2400 Bernville Road Reading, PA 19605 USA



IR-45563-100B (5/16)

¹ Licensed under US Patent Nos. 6,706,024; 6,872,195 and 7,329,412.

² A registered trademark of C. R. Bard, Inc.