



ARROW® PICC powered by ARROW® VPS TipTracker™ Stylet

REF **PR-45563-TTS**

6 Fr. | **3** Lumen | **55** cm catheter length | **.018** inch dia. spring-wire guide

Contents:

- 1: Multi-Lumen Indwelling TaperFree™ CG+™ Antimicrobial/Antithrombogenic Catheter¹: 6 Fr. x 55 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps, Contamination Guard, Arrow® VPS TipTracker™ Stylet and T-Port Connector
- 1: GlideThru™ Peel-Away Sheath: 6 Fr. x 2-3/4" (7 cm) Radiopaque over 6 Fr. Dilator
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer
- 1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 10 mL Luer-Lock
- 3: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SharpsAway® Disposal Cup

- 1: SharpsAway II™ Locking Disposal Cup
- 1: Catheter Trimmer
- 1: Safety Scalpel: #11
- 1: Patient Information Booklet
- 2: Paper Tape Measure
- 1: Tourniquet
- 1: Foam Electrodes (3 per pack)
- 1: Dressing: STATLOCK®² Catheter Stabilization Device
- 1: Equipment Cover: 13 cm x 244 cm (5" x 96")
- 1: T-Piece Cover

¹Licensed under US Patent Nos. 6,706,024, 6,872,195 and 7,329,412.
²A registered trademark of C. R. Bard, Inc.

All components are CE 0086 unless otherwise noted.

- nl** **CG+ ARROW PICC aangedreven door ARROW VPS TipTracker stilet**
Bewaren beneden 25°C. Vermijd hitte hoger dan 30°C.
- fr** **CCIP CG+ ARROW alimenté par le stylet ARROW VPS TipTracker**
Conservez le produit à une température inférieure à 25°C. Veillez à ne pas exposer le produit à une chaleur excessive supérieure à 30°C.
- de** **CG+ ARROW PEZK mit Antrieb durch ARROW VPS TipTracker Mandrin**
Unter 25°C lagern. Zu hohe Temperaturen über 30°C vermeiden.
- it** **PICC CG+ ARROW con mandrino VPS TipTracker ARROW**
Conservare ad una temperatura inferiore a 25°C. Evitare il calore eccessivo superiore a 30°C.
- pl** **Cewnik PICC CG+ ARROW zasilany mandrynem VPS TipTracker ARROW**
Przechowywać w temperaturze poniżej 25°C. Chronić przed temperaturą powyżej 30°C.
- pt** **PICC CG+ ARROW alimentado pelo estilete VPS TipTracker ARROW**
Conservar a temperaturas inferiores a 25°C. Evitar temperaturas superiores a 30°C.
- es** **Catéter central de inserción periférica CG+ ARROW alimentado por el estilete VPS TipTracker de ARROW**
Almacenar a temperaturas inferiores a 25°C. Evite temperaturas excesivas por encima de 30°C.
- sv** **CG+ ARROW perifert införd centralkateter drivs av ARROW VPS TipTracker-mandräng**
Förvaras vid temperatur under 25°C. Undvik höga temperatur över 30°C.


Contraindications: Clinical assessment of the patient must be completed to ensure no contraindications exist. The Arrow PICC with Chloragard® Technology is contraindicated in the following areas: Patients with known hypersensitivity to chlorhexidine; In presence of device related infections; In presence of previous or current thrombosis in the intended vessel or along the catheterized vessel pathway.



Not made with natural rubber latex



Store below 25°C (77°F). Avoid excessive heat above 30°C (86°F).



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.61	1060	5360	6
Medial (19 Ga.)	0.45	370	2090	not rated
Proximal (19 Ga.)	0.47	390	2160	not rated

* Priming volumes are approximate and are done without accessories.
 † Flow rates were determined using room temperature water, 100 cm head height and represent approximate flow capabilities.
 ‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.
 ** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.