



## Arterial Catheterization Kit

**REF SAC-00520-1A** 

**20**<sup>Ga</sup>

5 cm catheter length

.021

inch dia. spring-wire quide

## **Contents:**

- 1: Indwelling Catheter: 20 Ga. x 5 cm Radiopaque PEBA
- 1: Spring-Wire Guide, Marked: .021" (0.53 mm) dia. x 13-3/4" (35 cm) (Straight Soft Tip on Both Ends)
- 1: Injection Needle: SafetyGlide<sup>™1</sup> 25 Ga. x 1" (2.54 cm)
- 1: Introducer Needle: 20 Ga. x 1-1/2" (4 cm) XTW
- 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Syringe: 3 mL Luer-Lock
- 1: 5 mL 1% Lidocaine HCl Solution
- 1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®2 One-Step Solution with Hi-Lite Orange™ Tint (non-sterile solution)
- 1: SharpsAway® Disposal Cup
- 1: SharpsAway II™ Locking Disposal Cup
- 1: Drape: 36" x 41" (91 cm x 104 cm) with 3" x 5" (8 cm x 13 cm) fenestration
- 1: Towel: 13" x 18" (33 cm x 46 cm)
- 1: Needle Holder

- 1: Safety Scalpel: #11
- 1: Medication Label: 1% Lidocaine
- 4: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 4: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: Tegaderm<sup>®3</sup> 10 cm x 12 cm
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Suture: 3-0 Silk with Curved Needle
- <sup>1</sup>A trademark of Becton, Dickinson and Company.
- <sup>2</sup>A registered trademark of CareFusion or one of its subsidiaries.
- <sup>3</sup>A registered trademark of 3M Company.



















Not made with



Store between 20 - 25°C (68 - 77°F)

 $\mbox{\rm Rx}\,$  only. Warning: Read all package insert warnings, cautions and instructions prior to use. Failure to do so may result in severe patient injury or death.

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

🛕 California Prop. 65

Warning: Cancer and Reproductive Harm www.P65Warnings.ca.gov.

Fluid path components are non-pyrogenic.

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

Arrow International, Inc.
Subsidiary of Teleflex Incorporated
2400 Bernville Road
Reading, PA 19605 USA



IS-00520-101A (09/18)



Arterial Catheterization Kit

