



ARROW

ARROW® PICC Set

REF PS-01452-NM 4 Fr. | 2 Lumen | 50 cm catheter length

Contents:

- 1: Two-Lumen Indwelling Catheter: 4 Fr. x 50 cm Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps
- 1: Peel-Away Sheath: 15 Ga. x 1-1/4" (3.18 cm) Radiopaque over 17 Ga. TW Introducer Needle
- 2: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: Drape: 18" x 26" (46 cm x 66 cm) with 3" (7.6 cm) fenestration
- 1: Dressing: STATLOCK®¹ Catheter Stabilization Device

¹A registered trademark of C. R. Bard, Inc.

Rx only.

Warning: Read all package insert warnings, cautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of previous or current thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of patient must be completed to ensure no contraindications exist.

California Prop. 65
Warning: Cancer and Reproductive Harm
www.P65Warnings.ca.gov

Fluid path components are non-pyrogenic.

Not made with natural rubber latex

Catheter Cross Section

| Lumen | Priming Volume* (mL) | Flow Rate† (mL/hr) | Pump Flow Rate‡ (mL/hr) |
|-------------------|----------------------|--------------------|-------------------------|
| Distal (22 Ga.) | 0.27 | 250 | 1440 |
| Proximal (22 Ga.) | 0.33 | 210 | 1520 |

* Priming volumes are approximate and are done without accessories.
 † Flow rates are done with normal saline, room temperature, 100 cm head height and represent approximate flow capabilities.
 ‡ Pump flow rates determined at maximum pump pressure of 10 psig.

Manufacturer:
 Arrow International, Inc.
 Subsidiary of Teleflex Incorporated
 2400 Bernville Road
 Reading, PA 19605 USA

Teleflex®
www.teleflex.com

IS-01452-100D (3/18)



ARROW

ARROW® PICC Set

REF PS-01452-NM 4 Fr. | 2 Lumen | 50 cm catheter length

Contents:

- 1: Two-Lumen Indwelling Catheter: 4 Fr. x 50 cm Radiopaque Polyurethane with Blue FlexTip®, Extension Line Clamps
- 1: Peel-Away Sheath: 15 Ga. x 1-1/4" (3.18 cm) Radiopaque over 17 Ga. TW Introducer Needle
- 2: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: Drape: 18" x 26" (46 cm x 66 cm) with 3" (7.6 cm) fenestration
- 1: Dressing: STATLOCK®¹ Catheter Stabilization Device

¹A registered trademark of C. R. Bard, Inc.

Rx only.

Warning: Read all package insert warnings, cautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

Contraindications: The PICC is contraindicated wherever there is presence of device related infections or presence of previous or current thrombosis in the intended insertion vessel or catheter pathway. Clinical assessment of patient must be completed to ensure no contraindications exist.

California Prop. 65
Warning: Cancer and Reproductive Harm
www.P65Warnings.ca.gov

Fluid path components are non-pyrogenic.

Not made with natural rubber latex

Catheter Cross Section

| Lumen | Priming Volume* (mL) | Flow Rate† (mL/hr) | Pump Flow Rate‡ (mL/hr) |
|-------------------|----------------------|--------------------|-------------------------|
| Distal (22 Ga.) | 0.27 | 250 | 1440 |
| Proximal (22 Ga.) | 0.33 | 210 | 1520 |

* Priming volumes are approximate and are done without accessories.
 † Flow rates are done with normal saline, room temperature, 100 cm head height and represent approximate flow capabilities.
 ‡ Pump flow rates determined at maximum pump pressure of 10 psig.

Manufacturer:
 Arrow International, Inc.
 Subsidiary of Teleflex Incorporated
 2400 Bernville Road
 Reading, PA 19605 USA

Teleflex®
www.teleflex.com

IS-01452-100D (3/18)



ARROW® PICC Set



REF PS-01452-NM 4 Fr. | 2 Lumen | 50 cm catheter length



ARROW® PICC Set



REF PS-01452-NM 4 Fr. | 2 Lumen | 50 cm catheter length