



ARROW® PICC powered by ARROW® VPS Precision™ Stylet



REF **CDA-44041-VPS2**

4.5 Fr. | **40** cm catheter length | **.018** inch dia. spring-guide

ErgoPack®

Contents:

- 1: Indwelling Catheter with Chloragard Technology¹ and TaperFree® Design: 4.5 Fr. x 40 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip® Feature, Extension Line Clamps, Contamination Guard, Arrow® VPS Precision™ Stylet, Tuohy-Borst Adapter² and Marking Device
- 1: GlideThru™ Peel-Away Sheath: 4.5 Fr. x 2-3/4" (7 cm) Radiopaque over 4.5 Fr. Dilator
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer
- 1: Injection Needle: SafetyGlide™³ 25 Ga. x 5/8" (1.60 cm) RW
- 1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 10 mL Luer-Lock
- 1: Syringe: 3 mL Luer-Lock
- 1: Skin Protectant Prep Pad (non-sterile solution)
- 1: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SharpsAway® Disposal Cup

- 1: SharpsAway II™ Locking Disposal Cup
- 1: Catheter Trimmer
- 1: Towel: 24" x 36" (61 cm x 91 cm)
- 1: Drape: 60" x 76" (152 cm x 193 cm)
- 1: Drape: 36" x 41" (91 cm x 104 cm) with 3" x 5" (8 cm x 13 cm) fenestration
- 1: Particulate Filter: 5 Micron Straw
- 1: Safety Scalpel: #11
- 1: Sterile Procedure Sign
- 1: Patient ID Card
- 1: Chart Sticker
- 1: Patient Information Booklet
- 1: Checklist/CLIP Sheet
- 1: Flow Rate Information Card with Injection Log
- 2: Paper Tape Measure
- 1: Tourniquet
- 1: Foam Electrodes (3 per pack)
- 10: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 10: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: Tegaderm®⁴ 15.5 cm x 10 cm

- 1: Surgical Apparel: Impervious Gown
- 1: Dressing: STATLOCK®⁴ Catheter Stabilization Device
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask
- 1: Surgical Tape: Transpore™ 1" x 54" Roll
- 1: HemoHopper™ Fluid Receptacle
- 5: Console Pointers

¹Licensed under US Patent Nos. 6,706,024, 6,872,195 and 7,329,412.
²A trademark of Becton, Dickinson and Company.
³A registered trademark of 3M Company.
⁴A registered trademark of C. R. Bard, Inc.

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Contraindications: Clinical assessment of the patient must be completed to ensure no contraindications exist. The Arrow Chloragard PICC is contraindicated in the following areas: • Patients with known hypersensitivity to chlorhexidine • In presence of device related infections • In presence of previous or current thrombosis in the intended vessel or along the catheterized vessel pathway

Warning: The results of certain animal experiments have shown phthalates to be potentially toxic to reproduction. Proceeding from the present state of scientific knowledge, risks for male premature infants cannot be excluded in the case of long-term exposure or application. Medical products containing phthalates should be used only temporarily with pregnant women, nursing mothers, babies and infants.




Not made with natural rubber latex



⚠️ Precaution: Contains Phthalate: DEHP



Store below 25°C (77°F). Avoid excessive heat above 30°C (86°F).



Lumen	Priming Volume* (mL)	Flow Rate† (mL/hr)	Pump Flow Rate‡ (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (17 Ga.)	0.49	1510	6350	5

* Priming volumes are approximate and are done without accessories.
 † Flow rates were determined using room temperature water, 100 cm head height and represent approximate flow capabilities.
 ‡ Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.
 **Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.