



ARROW® PICC with Chloragard® Technology

REF CDA-44063-HPK1A

ErgoPack®

Contents:


- 1: Multi-Lumen Indwelling TaperFree™ CG+™ Antimicrobial/Antithrombogenic Catheter¹: 6 Fr. x 40 cm Pressure Injectable Radiopaque Polyurethane with Blue FlexTip®, T-Port Connector, Extension Line Clamps, Contamination Guard and Placement Wire
- 1: GlideThru™ Peel-Away Sheath: 6 Fr. x 2-3/4" (7 cm) Radiopaque over 6 Fr. Dilator
- 1: Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - Straight Stiff Tip on Other) with Arrow Advancer™
- 1: Injection Needle: 25 Ga. x 1" (2.5 cm)
- 1: Safety Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Introducer Needle: Echogenic 21 Ga. x 2-3/4" (7 cm) TW
- 1: Syringe: 10 mL Luer-Lock
- 1: Syringe: 3 mL Luer-Lock
- 3: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SharpsAway II™ Locking Disposal Cup
- 1: Catheter Trimmer
- 1: Drape: 24" x 36"
- 1: Drape: 60" x 76"
- 1: Drape: 36" x 41" with 3" x 5" fenestration
- 1: Towel
- 1: Filter: 5 Micron Straw
- 1: Safety Scalpel: #11
- 1: Patient ID Card
- 1: Patient Information Booklet
- 2: Paper Tape Measure
- 1: Tourniquet
- 10: Gauze Pad: 2" x 2"
- 10: Gauze Pad: 4" x 4"
- 1: Surgical Apparel: Mask
- 1: Dressing: Tegaderm®² 15.5 cm x 10 cm
- 1: Surgical Apparel: Impervious Gown
- 1: Dressing: STATLOCK®³ Catheter Stabilization Device
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Surgical Apparel: Bouffant Cap
- 1: Tape: Steri-Strip®²
- 1: HemoHopper® Fluid Receptacle

¹ Licensed under US Patent Nos. 6,706,024; 6,872,195 and 7,329,412.

² A registered trademark of 3M Company.

³ A registered trademark of C. R. Bard, Inc.

All components are CE 0086 unless otherwise noted.



Catheter Cross Section

Lumen	Priming Volume* (mL)	Flow Rate [†] (mL/hr)	Pump Flow Rate [‡] (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (18 Ga.)	0.50	1370	6970	6
Medial (19 Ga.)	0.39	490	2810	not rated
Proximal (19 Ga.)	0.41	510	2880	not rated

* Priming volumes are approximate and are done without injection caps.

[†] Flow rates were determined using room temperature water, 100 cm head height and represent approximate flow capabilities.

[‡] Pump flow rates are determined at maximum pump pressure of 10 psig and represent approximate flow capabilities.

** Pressure injection flow rates are determined at the injector pressure setting of 300 psi maximum using media of 11.8 centipoise viscosity, with 152 cm pressure tubing.

- nl** ARROW PICC met Chloragard-technologie
Bewaren beneden 25°C. Vermijd hitte hoger dan 30°C.
- fr** PICC ARROW avec technologie Chloragard
Conservez le produit à une température inférieure à 25°C. Veillez à ne pas exposer le produit à une chaleur excessive supérieure à 30°C.
- de** ARROW PICC mit Chloragard-Technologie
Unter 25°C lagern. Zu hohe Temperaturen über 30°C vermeiden.
- el** Καθετήρας ARROW PICC με Τεχνολογία Chloragard
Φυλάσσεται κάτω των 25°C. Να αποφεύγεται η υπερβολική θερμότητα άνω των 30°C.
- it** ARROW PICC con tecnologia Chloragard
Conservare ad una temperatura inferiore a 25°C. Evitare il calore eccessivo superiore a 30°C.
- pl** Cewnik ARROW PICC z technologią Chloragard
Przechowywać w temperaturze poniżej 25°C. Chronić przed temperaturą powyżej 30°C.
- pt** PICC ARROW com Tecnologia Chloragard
Conservar a temperaturas inferiores a 25°C. Evitar temperaturas superiores a 30°C.
- ru** ARROW PICC с технологией Chloragard
Хранить при температуре ниже 25°C. Избегать нагрева свыше 30°C.
- es** PICC de ARROW con tecnología Chloragard
Almacenar a temperaturas inferiores a 25°C. Evite temperaturas excesivas por encima de 30°C.
- sv** ARROW PICC med Chloragard-teknologi
Förvaras vid temperatur under 25°C. Undvik höga temperatur över 30°C.
- tr** Chloragard Teknolojili ARROW PICC
25°C'ın altında saklayın. 30°C'ın üzerindeki aşırı ısıdan sakının.