





20 cm catheter length 1032 inch dia. spring-wire guide

## Two-Lumen Central Venous Catheterization Kit with Blue FlexTip® ARROWg+ard Blue® Catheter

## Contents:

- Arrowg+ard Blue® Two-Lumen Catheter: 7 Fr. (2.5 mm OD) x 20 cm Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow Advancer Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle

- Needle
  Pressure Transduction Probe
  Injection Needle: 25 Ga. x 1" (2.54 cm) and 3 mL Luer-Lock Syringe
  Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow®
  Raulerson Spring-Wire Introduction Syringe
  Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
  Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm
  Dust Cap: Non-Vented
  SecondSite™ Adjustable Hub: Fastener
  SecondSite™ Adjustable Hub: Catheter Clamp
  SharpsAway® Disposal Cup
  Drape: 21" x 36" (53 cm x 91 cm) with 3" (7.6 cm) fenestration, with adhesive
  Scalpel: #11

- 1: Scalpel: #11
  2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
  5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
  1: Suture: 3-0 Braided Silk with Straight Needle

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.
California Prop. 65

MWARNING: Cancer and Reproductive Harm.
www.P65Warnings.ca.gov
Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F).

Fluid path components are non-pyrogenic.

Contraindications: The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa drugs.

6 Lumen	Priming Volume* (mL)	Flow Rate† (mL/hr)	
Distal (16 Ga.)	0.45	2540	
Proximal (16 Ga.)	0.49	2888	

<sup>\*</sup> Priming volumes are approximate and are done without accessories.




















Arrow International, Inc. Subsidiary of Teleflex Incorporated 2400 Bernville Road Reading, PA 19605 USA



LBL026708 R01 (2021-11)









 $<sup>\</sup>ensuremath{\dagger}$  Flow rate values are approximate and are determined using deionized water at 100 cm head height.