



**REF SA-15853** 

8.5 Fr. 3 Lumen

20 cm catheter length

.032 inch dia. spring-wire guide

## **Three-Lumen CVC**

## Contents:

- Three-Lumen Catheter: 8.5 Fr. (3.0 mm OD) x 20 cm
- Spring-Wire Guide, Nitinol, Marked: .032" (0.81 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow Advancer with
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW

- Introducer Needle: Ecnogenic 18 Ga. x 2-1/2" (6.35 cm) X I W
  Pressure Transduction Probe
  Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL
  Tissue Dilator: 9 Fr. (3.0 mm) x 10.2 cm
  SecondSite ™ Adjustable Hub: Fastener
  SecondSite ™ Adjustable Hub: Catheter Clamp
  Drape: 29-1/2" x 35-1/2" (75 cm x 90 cm) with 4" (10 cm) fenestration, with
  adhesive adhesive
  Drape: 19-5/8" x 29-1/2" (50 cm x 75 cm) with adhesive
- Safety Scalpel: #11
- Prep Sponge Swab

  Valve: MicroCLAVE™¹ Neutral Displacement Connector
- <sup>1</sup>A trademark of ICU Medical, Inc.

All components are CE 2797 unless otherwise noted.

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU Not made with natural rubber latex. Fluid path components are non-pyrogenic.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
Distal (16 Ga.)	0.47	2930	
Medial (14 Ga.)	0.89	6267	
Proximal (16 Ga.)	0.47	3731	

<sup>\*</sup> Priming volumes are approximate and are done without accessories.



LOT



M













**EU Authorized** Representative and Importer



Teleflex Medical IDA Business and Technology Park Dublin Road, Athlone Co. Westmeath, Ireland

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL050082 R01 (2023-09)









<sup>†</sup> Flow rate values are approximate and are determined using deionized water at 100 cm head height.