





REF AK-22123-CDC

Arrowg+ard Blue® Large-Bore Three-Lumen CVC for High Volume Infusions

Contents:

- Arrowg+ard Blue® Three-Lumen Catheter: 12 Fr. (4.1 mm OD) x 16 cm Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow Advancer
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- Injection Needle: 25 Ga. x 1" (2.54 cm)
 Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- Pressure Transduction Probe
- Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
- Syringe: 3 mL Luer-Lock
 Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm
 5 mL 1% Lidocaine HCl Solution and Alcohol Prep
- o III. 176 Lidocaine HCI Solution and Alcohol Prep
 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®¹ One-Step Solution with Hi-Lite Orange™ Tint
 SharpsAway® II Locking Disposal Cup
 SharpsAway® Disposal Cup
 Drape: 41" x 55" with 4" fenestration
 Needle Holder
 Softh Solution #44

- Safety Scalpel: #11
 Gauze Pad: 2" x 2" (5 cm x 5 cm)
 Gauze Pad: 4" x 4" (10 cm x 10 cm)
- Suture: 3-0 Silk with Curved Needle
- ¹A registered trademark of CareFusion or one of its subsidiaries.

	Х	10	1	l
			٠	

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. California Prop. 65

MARNING: Cancer and Reproductive Harm.

www.P65Warnings.ca.gov

Not made with natural rubber latex.

Store between 20 - 25°C (68 - 77°F) Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic

Contraindications: The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa drugs.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
Distal (16 Ga.)	0.35	2581	
Medial (12 Ga.)	1.2	9719	
Proximal (12 Ga.)	1.1	12666	

Priming volumes are approximate and are done without accessories.









Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL050972 R01 (2022-02)











 $[\]ensuremath{\dagger}$ Flow rate values are approximate and are determined using deionized water at 100 cm head height.