



**REF SA-15553** 

5.5<sup>Fr.</sup>

3<sup>Lumen</sup>





**.018** inch dia. spring-wire guide

## **Three-Lumen CVC**

## Contents:

- Three-Lumen Catheter: 5.5 Fr. (1.9 mm OD) x 8 cm 1.
- Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow Advancer 1:
- Catheter: 22 Ga. x 1-3/4" (4.45 cm) Radiopaque over 25 Ga. RW Introducer Needle 1:
- Introducer Needle: 20 Ga. x 1-1/2" (4 cm) XTW
- 1.
- 1:
- 1:
- Syringe: 5 mL Luer-Slip Tissue Dilator: 6 Fr. (2.0 mm) x 7.6 cm Tissue Dilator: 6 Fr. (2.0 mm) x 4.1 cm SecondSite™ Adjustable Hub: Fastener SecondSite™ Adjustable Hub: Catheter Clamp Cefot: Concelst.#11 1: 1:
- Safety Scalpel: #11 Valve: MicroCLAVE<sup>TM1</sup> Neutral Displacement Connector 1.
- 3:
- <sup>1</sup>A trademark of ICU Medical, Inc. All components are CE 2797 unless otherwise noted.

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU Not made with natural rubber latex.

Fluid path components are non-pyrogenic.



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
Distal (20 Ga.)	0.29	1253	
Medial (22 Ga.)	0.31	601	
Proximal (22 Ga.)	0.34	872	

\* Priming volumes are approximate and are done without accessories.

+ Flow rate values are approximate and are determined using deionized water at 100 cm head height.



EU Authorized Representative and Importer:

IDA Business and E Technology Park Dublin Road, Athlone 2797 Co. Westmeath, Ireland

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL051287 R01 (2023-09)











**Three-Lumen CVC** 



