



**REF CA-24301 16** Ga. **1** Lumen



# Arrowg+ard Blue® One-Lumen CVC

### Contents

- 1: Arrowg+ard Blue® One-Lumen Catheter: 16 Ga. (1.8 mm OD) x 20
- Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow 1: Advancer
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW 1: Introducer Needle
- 1: Pressure Transduction Probe
- Injection Needle: 25 Ga. x 1" (2.54 cm) and 3 mL Luer-Lock Syringe Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow®
- 1: Raulerson Spring-Wire Introduction Syringe Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip
- 1:
- Syringe Tissue Dilator: 7 Fr. (2.3 mm) x 10.2 cm 1:
- 1:
- Ilssue Dilator, r r (2,5 min) > 10,5 min) 1:
- SharpsAway® Disposal Cup Drape: 21" x 36" (53 cm x 91 cm) with 3" (7.6 cm) fenestration, with 1: adhesive

### Rx only

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F).

Sterilized using ethylene oxide unless otherwise indicated in the contents list. Fluid path components are non-pyrogenic.

Contraindications: The Arrowg+ard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/ or sulfa drugs.

<b>O</b> Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
Distal (16 Ga.)	0.44	2781	

\* Priming volumes are approximate and are done without accessories. † Flow rate values are approximate and are determined using deionized water at 100 cm head height.

## 

Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL052071 R00 (2021-07)



LOT

**REF CA-24301** 

~~~|







## Arrowg+ard Blue<sup>®</sup> One-Lumen CVC

1: Scalpel: #11 3: Gauze Pad: 4" x 4" (10 cm x 10 cm)

1: Suture: 3-0 Braided Silk with Straight Needle