



REF AK-14703-SP

7 Fr. 3 Lumen 30 cm catheter length .032 inch dia. spring-wire guide

# Three-Lumen CVC

**Contents:**

- 1: Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 30 cm
- 1: Spring-Wire Guide: .032" (0.81 mm) dia. x 26-3/4" (68 cm) (Straight Soft Tip on One End - "J" Tip on Other)
- 1: Injection Needle: 25 Ga. x 1" (2.54 cm)
- 1: Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW
- 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle with 5 mL Luer-Slip Syringe
- 1: Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe
- 1: Syringe: 3 mL Luer-Lock
- 1: Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm
- 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep
- 1: 3 mL Applicator 2% CHG and 70% IPA Chloraprep<sup>®1</sup> One-Step Solution with Hi-Lite Orange™ Tint
- 3: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SharpsAway® II Locking Disposal Cup
- 1: SharpsAway® Disposal Cup
- 1: Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration
- 1: Safety Scalpel: #11
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Suture: 3-0 Braided Silk with Straight Needle

<sup>1</sup>A registered trademark of CareFusion or one of its subsidiaries.

**Rx only**

**Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.**

**California Prop. 65**

**⚠ WARNING:** Cancer and Reproductive Harm. [www.P65Warnings.ca.gov](http://www.P65Warnings.ca.gov)

**Not made with natural rubber latex.**

Store between 20 – 25°C (68 – 77°F).

Sterilized using ethylene oxide unless otherwise indicated in the contents list.

Fluid path components are non-pyrogenic.



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (16 Ga.)	0.54	2025
Medial (18 Ga.)	0.45	896
Proximal (18 Ga.)	0.48	955

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.



LOT



Arrow International LLC  
Subsidiary of Teleflex Incorporated  
3015 Carrington Mill Blvd.  
Morrisville, NC 27560 USA

**Teleflex**  
[www.teleflex.com](http://www.teleflex.com)

LBL052779 R01 (2022-03)



REF AK-14703-SP

7 Fr. 3 Lumen 30 cm catheter length .032 inch dia. spring-wire guide

# Three-Lumen CVC