



REF CDC-22123-X1A (IPNXXXXXX)

12 Fr. 3 Lumen 16 cm catheter length .035 inch dia. spring-wire guide



Arrowgard Blue® Large-Bore Three-Lumen CVC for High Volume Infusions



Contents:

- | | | |
|--|--|---|
| <ul style="list-style-type: none"> 1: Arrowgard Blue® Three-Lumen Catheter: 12 Fr. (4.1 mm OD) x 16 cm
Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Injection Needle: SafetyGlide™ 1 25 Ga. x 1" (2.54 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Pressure Transduction Probe Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Injection Needle: SafetyGlide™ 1 23 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Syringe: 3 mL Luer-Lock Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Tissue Dilator: 10 Fr. (3.5 mm) x 10.2 cm Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 2: 10 mL Luer-Lock Pre-Filled Saline Syringe Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: 5 mL 1% Lidocaine HCl Solution and Alcohol Prep Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX | <ul style="list-style-type: none"> 1: 3 mL Applicator 2% CHG and 70% IPA ChloraPrep® 2 One-Step Solution with Hi-Lite Orange™ Tint Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: SharpsAway® II Locking Disposal Cup Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: SharpsAway® Disposal Cup Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Maximal Barrier Drape™ with 4" fenestration Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Towel Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Needle Holder Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Safety Scalpel: #11 Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Checklist/CLIP Sheet Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Sterile Procedure Sign Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Medication Label: 1% Lidocaine Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 2: Gauze Pad: 2" x 2" (5 cm x 5 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 5: Gauze Pad: 4" x 4" (10 cm x 10 cm) Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Dressing: Tegaderm® 3 10 cm x 12 cm Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Surgical Apparel: Impervious Gown Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Dressing: BIOPATCH® 4 Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Surgical Apparel: Bouffant Cap Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX | <ul style="list-style-type: none"> 1: Surgical Apparel: Mask with Eye Shield Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: Suture: 3-0 Silk with Curved Needle Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX 1: HemoHopper® Fluid Receptacle Origin: XX XXXX XX XXX XXXXXXXX XXXXXXXXXX <p> ¹A trademark of Becton, Dickinson and Company.
 ²A registered trademark of CareFusion or one of its subsidiaries.
 ³A registered trademark of 3M Company.
 ⁴A registered trademark of Johnson & Johnson Corporation. </p> |
|--|--|---|

Not made with natural rubber latex.

Store between 20 – 25°C (68 – 77°F).

Arrow International provides the enclosed medication label(s) for your convenience. Please ensure that these labels are applied to the correct syringe and corresponding medication.

Contraindications: The Arrowgard Blue antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa drugs.



Lumen

	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (16 Ga.)	0.35	2581
Medial (12 Ga.)	1.2	9719
Proximal (12 Ga.)	1.1	12666

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

