







(IPNXXXXXX)





Pressure Injectable Arrowg+ard Blue Plus® Four-Lumen CVC Kit



Contents:

- 1: Arrowg+ard Blue Plus®1 Four-Lumen Catheter: 8.5 Fr. (3.0 mm OD) x 16 cm, Pressure Injectable
- Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- Pressure Transduction Probe
- Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- Injection Needle: Eclipse™2 23 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-
- Injection Needle: Eclipse™2 25 Ga. x 1" (2.54 cm) and 3 mL Luer-Lock Syringe
- Syringe: 10 mL Luer-Lock
- Tissue Dilator: 10 Fr. (3.5 mm) x 10.2 cm
- SecondSite™ Adjustable Hub: Fastener
- SecondSite™ Adjustable Hub: Catheter Clamp
- SharpsAway® II Locking Disposal Cup
- SharpsAway® Disposal Cup
- Maximal Barrier Drape™ with 4" fenestration
- Needle Holder
- Forceps
- Safety Scalpel: #11
- Checklist/CLIP Sheet

- Sterile Procedure Sign
- Flow Rate Information Card with Injection Log
- Prep Sponge Swab
- Gauze Pad: 2" x 2" (5 cm x 5 cm)
- Gauze Pad: 4" x 4" (10 cm x 10 cm)
- Dressing: Tegaderm®3 10 cm x 12 cm
 - Suture: 3-0 Silk with Curved Needle
- 1: HemoHopper® Fluid Receptacle

¹Licensed under US Patent No. 7.329.412.

²A trademark of Becton, Dickinson and Company.

3A registered trademark of 3M Company.



Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F)

Contraindications: The Arrowg+ard Blue Plus antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine. silver sulfadiazine and/or sulfa drugs

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (16 Ga.)	0.38	2781	5
Medial 1 (14 Ga.)	0.45	4980	10
Medial 2 (18 Ga.)	0.35	1453	not rated
Proximal (18 Ga.)	0.38	1619	not rated

^{*} Priming volumes are approximate and are done without accessories





LBL053370 R02 (2024-04)

[†] Flow rate values are approximate and are determined using deionized water at 100 cm head

^{**} Pressure injection flow rates are determined using media of 11.8 centipoise (cP) viscosity, with