







Contraindications: The Arrowg+ard Blue Plus antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine.

(IPNXXXXXX)

■032 inch dia. spring-wire guide

Pressure Injectable Arrowg+ard Blue Plus® Two-Lumen CVC Kit



Contents:

- Arrowg+ard Blue Plus®1 Two-Lumen Catheter: 8 Fr. (2.8 mm OD) x 16 cm. Pressure Injectable
- Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- Pressure Transduction Probe
- Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- Injection Needle: SafetyGlide™2 25 Ga. x 1" (2.54 cm) and 3 mL Luer-
- Injection Needle: SafetyGlide™2 23 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe

Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F)

- Syringe: 10 mL Luer-Lock
- Tissue Dilator: 9 Fr. (3.0 mm) x 10.2 cm
- SecondSite™ Adjustable Hub: Fastener
- SecondSite™ Adjustable Hub: Catheter Clamp
- SharpsAway® II Locking Disposal Cup
- SharpsAway® Disposal Cup
- Maximal Barrier Drape™ with 4" fenestration
- Safety Scalpel: #11
- Checklist/CLIP Sheet
- Sterile Procedure Sign
- Flow Rate Information Card with Injection Log

Not made with natural rubber latex.

silver sulfadiazine and/or sulfa drugs.

- Gauze Pad: 2" x 2" (5 cm x 5 cm)
- Gauze Pad: 4" x 4" (10 cm x 10 cm)
- Dressing: Tegaderm®3 CHG 3-1/2" x 4-1/2" (8.5 cm x 11.5 cm)
- Suture: 3-0 Silk with Curved Needle
- 2: Valve: MaxZero™4 Needleless Connector
- 1: HemoHopper® Fluid Receptacle
- ¹Licensed under US Patent No. 7,329,412.
- ²A trademark of Becton, Dickinson and Company.
- 3A registered trademark of 3M Company.
- ⁴A trademark of CareFusion or one of its subsidiaries



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	MAX Pressure Injection Flow Rate** (mL/sec)
Distal (14 Ga.)	0.68	6350	10
Proximal (14 Ga.)	0.70	4980	10

^{*} Priming volumes are approximate and are done without accessories.



LBL053382 R01 (2024-01)

[†] Flow rate values are approximate and are determined using deionized water at 100 cm head

^{**} Pressure injection flow rates are determined using media of 11.8 centipoise (cP) viscosity, with