

(IPNXXXXXX)

7^{Fr.} **3**^{Lum}

16 cm catheter length

.032 inch dia. spring-wire guide



Pressure Injectable Arrowg+ard Blue Plus® Three-Lumen CVC



Contents:

- Arrowg+ard Blue Plus®¹ Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 16 cm, Pressure Injectable
- 1: Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow Advancer
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- Pressure Transduction Probe
- Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- Injection Needle: Eclipse™2 22 Ga. x 1-1/2" (3.81 cm) RW and 5 mL Luer-Slip Syringe
- Injection Needle: Eclipse™2 25 Ga. x 1" (2.54 cm) RW and 3 mL Luer-Lock Syringe
- : Syringe: 10 mL Luer-Lock
- 1: Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm
- 1: SecondSite™ Adjustable Hub: Fastener
- SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SharpsAway® II Locking Disposal Cup
- 1: SharpsAway® Disposal Cup
- 1: Maximal Barrier Drape™ with 4" fenestration
- 1. Towe
- 1: Needle Holder
- 1: Forceps
- 1: Safety Scalpel: #11
- 1: Checklist/CLIP Sheet
- 1: Sterile Procedure Sign
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)

- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: Tegaderm®3 10 cm x 12 cm
- : Surgical Apparel: Impervious Gown
- I: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Suture: 3-0 Silk with Curved Needle
- HemoHopper® Fluid Receptacle

¹Licensed under US Patent No. 7,329,412. ²A trademark of Becton, Dickinson and Company

³A registered trademark of 3M Company.










Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	MAX Pressu Injection Flow Rate* (mL/sec)
Distal (16 Ga.)	0.38	2490	10
Medial (18 Ga.)	0.37	1162	5
Dravimal (19 Ca)	0.30	1328	5

^{*} Priming volumes are approximate and are done without accessories.



Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F).

Contraindications: The Arrowg+ard Blue Plus antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa drugs.





LBL053592 R01 (2024-05)

 $[\]uparrow$ Flow rate values are approximate and are determined using deionized water at 100 cm head height.

^{**} Pressure injection flow rates are determined using media of 11.8 centipoise (cP) viscosity, with 152 cm administration set tubing.