



CDA-42802-P1A

(IPNXXXXXX)

8 Fr.

2 Lumen

16 cm catheter length

.032 inch dia. spring-wire guide

ARROW®

ErgoPack®

Pressure Injectable Arrowg+ard Blue Plus® Two-Lumen CVC



Contents:

- 1: Arrowg+ard Blue Plus®¹ Two-Lumen Catheter: 8 Fr. (2.8 mm OD) x 16 cm, Pressure Injectable
- 1: Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer
- 1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- 1: Pressure Transduction Probe
- 1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction Syringe
- 1: Injection Needle: Eclipse™² 22 Ga. x 1-1/2" (3.81 cm) RW and 5 mL Luer-Slip Syringe
- 1: Injection Needle: Eclipse™² 25 Ga. x 1" (2.54 cm) RW and 3 mL Luer-Lock Syringe
- 1: Syringe: 10 mL Luer-Lock
- 1: Tissue Dilator: 9 Fr. (3.0 mm) x 10.2 cm
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: SharpsAway® II Locking Disposal Cup
- 1: SharpsAway® Disposal Cup
- 1: Maximal Barrier Drape™ with 4" fenestration
- 1: Towel
- 1: Needle Holder
- 1: Forceps
- 1: Safety Scalpel: #11
- 1: Checklist/CLIP Sheet
- 1: Sterile Procedure Sign
- 2: Gauze Pad: 2" x 2" (5 cm x 5 cm)
- 5: Gauze Pad: 4" x 4" (10 cm x 10 cm)
- 1: Dressing: Tegaderm®³ 10 cm x 12 cm
- 1: Surgical Apparel: Impervious Gown
- 1: Surgical Apparel: Bouffant Cap
- 1: Surgical Apparel: Mask with Eye Shield
- 1: Suture: 3-0 Silk with Curved Needle
- 1: HemoHopper® Fluid Receptacle

¹Licensed under US Patent No. 7,329,412.

²A trademark of Becton, Dickinson and Company.

³A registered trademark of 3M Company.

en pt-BR es-MX



Lumen

Priming
Volume*
(mL)

Gravity
Flow
Rate†
(mL/hr)

MAX Pressure
Injection
Flow Rate**
(mL/sec)

Distal (14 Ga.)
Proximal (14 Ga.)

0.68
0.70

6350
4980

10
10

Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F).

Contraindications: The Arrowg+ard Blue Plus antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa drugs.

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

** Pressure injection flow rates are determined using media of 11.8 centipoise (cP) viscosity, with 152 cm administration set tubing.



LATEX



25°C
(77°F)