







16 cm catheter length .032 inch dia. spring-wire guide

Pressure Injectable Arrowg+ard Blue Plus® Three-Lumen CVC Kit

Contents:

- Arrowg+ard Blue Plus®1 Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 16 cm,
 Pressure Injectable
 Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight
 Soft Tip on One End "J" Tip on Other) with Arrow Advancer
 Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer
 Needle

- Needle
 Pressure Transduction Probe
 Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL
 Arrow® Raulerson Spring-Wire Introduction Syringe
 Injection Needle: Eclipse™ 2 3 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip
- Syringe Injection Needle: Eclipse $^{\text{TM}\,2}$ 25 Ga. x 1" (2.54 cm) and 3 mL Luer-Lock Syringe
 Tissue Dilator: 8.5 Fr. (2.8 mm) x 10.2 cm
 SecondSite™ Adjustable Hub: Fastener
 SecondSite™ Adjustable Hub: Catheter Clamp
 SharpsAway® II Locking Disposal Cup
 SharpsAway® Disposal Cup
 Drape: 21" x 36" (53 cm x 91 cm) with 3" (7.6 cm) fenestration, with adhesive
 Safety Scalpel: #11

Rx only
Warning: Read all package insert warnings, precautions, and
instructions prior to use. Failure to do so may result in severe patient
injury or death.
California Prop. 65

Injury or death.

California Prop. 65

MARNING: Cancer and Reproductive Harm.
www.P65Warnings.ca.gov

Not made with natural rubber latex.

Store below 25°C (77°F). Avoid excessive heat above 40°C (104°F).

Sterilized using ethylene oxide unless otherwise indicated in the contents

Contraindications: The Arrowg+ard Blue Plus antimicrobial catheter is contraindicated for patients with known hypersensitivity to chlorhexidine, silver sulfadiazine and/or sulfa drugs.

2:	Gauze	Pad:	2"	x 2"	(5	cm	X 5	cm)
5.	Gaura	Pad-	4"	ν 4"	(1)	n cn	n v	10 6	٠,

Lumen	Priming Flow Volume* Rate† (mL) (mL/hr)		MAX Pressure Injection Flow Rate** (mL/sec)	
Distal (16 Ga.)	0.38	2490	10	
Medial (18 Ga.)	0.37	1162	5	
Proximal (18 Ga.)	0.39	1328	5	







Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL053674 R02 (2023-12)











^{*} Priming volumes are approximate and are done without accessories.
† Flow rate values are approximate and are determined using deionized water at 100 cm head height.
** Pressure injection flow rates are determined using media of 11.8 centipoise (cP) viscosity, with 152 cm administration set tubing.