



REF AK-12123-F

16 cm catheter length .035 inch dia. spring-wire guide

Large-Bore Three-Lumen CVC for High Volume Infusions

Contents:

- Three-Lumen Catheter: 12 Fr. (4.1 mm OD) x 16 cm Spring-Wire Guide, Marked: .035" (0.89 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End "J" Tip on Other) with Arrow Advancer
- Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle
- Injection Needle: 25 Ga. x 1" (2.54 cm)
 Blunt Fill Needle: 18 Ga. x 1-1/2" (3.81 cm)
- Pressure Transduction Probe
- Introducer Needle: 18 Ga. x 2-1/2" (6.35 cm) XTW and 5 mL Arrow® Raulerson Spring-Wire Introduction
- Injection Needle: 22 Ga. x 1-1/2" (3.81 cm) and 5 mL Luer-Slip Syringe

- Syringe: 3 mL Luer-Lock
 Tissue Dilator: 12 Fr. (4.0 mm) x 14 cm
 5 mL 1% Lidocaine HCl Solution and Alcohol Prep
- 3 mL Applicator 2% CHG and 70% IPA ChloraPrep®¹ One-Step Solution with Hi-Lite Orange™ Tint Dust Cap: Non-Vented

- Dust Cap: Non-Vented
 SharpsAway® Disposal Cup
 Drape: 24" x 36" (61 cm x 91 cm) with 4" (10 cm) fenestration
 Scalpel: #11
 Gauze Pad: 2" x 2" (5 cm x 5 cm)
 Gauze Pad: 4" x 4" (10 cm x 10 cm)
 Suture: 3-0 Braided Silk with Straight Needle

- ¹A registered trademark of CareFusion or one of its subsidiaries.

Rx	10	1	ŀ
MA	rn.	:	

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

California Prop. 65

MARNING: Cancer and Reproductive Harm.

www.P65Warnings.ca.gov

Not made with natural rubber latex. Store between 20 - 25°C (68 - 77°F)

Sterilized using ethylene oxide unless otherwise indicated in the contents list. Fluid path components are non-pyrogenic.

Θ	
Lumen	

Gravity Flow Rate† (mL/hr) Priming Volume* (mL)

Distal (16 Ga.) Medial (12 Ga.) Proximal (12 Ga.)

 \uparrow Flow rate values are approximate and are determined using deionized water at 100 cm head height.








Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL053675 R01 (2022-02)









^{*} Priming volumes are approximate and are done without accessories.