



## REF MP-15703-JJ

20 cm catheter length

.018 Inch ula. spring-wire guide

## Three-Lumen CVC

## Contents:

- Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 20 cm,
- Spring-Wire Guide, Nitinol, Marked: .018" (0.46 mm) dia. x 23-5/8" (60 cm) (Straight Stiff Tip on One End -
- "J" Tip on Other) with Arrow Advancer and hub Catheter: 20 Ga. x 1-3/4" (4.45 cm) Radiopaque over 22 Ga. TW Introducer Needle
- Injection Needle: 25 Ga. x 1-1/2" (3.81 cm)
- Injection Needle: 22 Ga. x 1-1/2" (3.81 cm)
- Introducer Needle: Echogenic 22 Ga. (0.72 mm) x
- 2-3/4" (7 cm) XTW Introducer Needle: Echogenic 22 Ga. (0.72 mm) x 1-3/8" (3.5 cm) XTW
- Syringe: 10 mL Luer-Slip
- Syringe: 5 mL Luer-Slip
- Syringe: 2 mL Luer-Slip
- 1. Tissue Dilator: 8 Fr. x 9.7 cm
- Dust Cap: Non-Vented
  SecondSite™ Adjustable Hub: Fastener
- SecondSite™ Adjustable Hub: Catheter Clamp
- Medicine Cup: 60 mL 1:
- SharpsAway® Disposal Cup
- Drape: 35" x 50" (89 cm x 127 cm) with 4" (10 cm)

- fenestration, with adhesive
- Needle Holder
- Scalpel: #11
- 2: Stopcock: 4-way
- 6:
- Prep Sponge Swab Gauze Pad: 4" x 4" (10 cm x 10 cm) Dressing: OpSite®1 4-3/4" x 4" (12 cm x 10 cm)
- Suture: 3-0 Silk with Curved Needle 1:
- Tray: Prep
- Y-hub Guidewire Introducer
- <sup>1</sup>A registered trademark of T.J. Smith & Nephew,

Limited.

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.

Not made with natural rubber latex.

Sterilized using ethylene oxide unless otherwise indicated in the contents

Fluid path components are non-pyrogenic.

Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)	
Distal (16 Ga.)	0.47	2656	
Medial (18 Ga.)	0.4	1245	
Proximal (18 Ga.)	0.42	1328	

<sup>\*</sup> Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head













Arrow International LLC Subsidiary of Teleflex Incorporated 3015 Carrington Mill Blvd. Morrisville, NC 27560 USA



LBL053749 R00 (2021-10)



