



REF MTO-12703-J

7 Fr. 3 Lumen 16 cm catheter length .032 inch dia. spring-wire guide

Three-Lumen CVC

Contents:

- 1: Three-Lumen Catheter: 7 Fr. (2.5 mm OD) x 16 cm

1: Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 17-3/4" (45 cm) (Straight Soft Tip on One End – "J" Tip on Other) with Arrow Advancer

1: Catheter: 18 Ga. x 2-1/2" (6.35 cm) Radiopaque over 20 Ga. RW Introducer Needle

1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW

1: Pressure Transduction Probe

1: Injection Needle: 23 Ga. x 1-1/4" (3.20 cm)

2: Syringe: 3 mL Luer-Slip

1: Syringe: 5 mL Luer-Slip

1: Arrow® Raulerson Spring-Wire Introduction Syringe: 5 mL

1: Tissue Dilator: 8 Fr. (2.7 mm) x 9 cm

3: Dust Cap: Non-Vented

1: SecondSite™ Adjustable Hub: Fastener

1: SecondSite™ Adjustable Hub: Catheter Clamp

1: SharpsAway® Disposal Cup

1: Drape: 35" x 50" (89 cm x 127 cm) with 4" (10 cm) fenestration, with adhesive

1: Needle Holder

1: Scalpel: #11

1: Dressing: Adhesive 5" x 4"(12.5 cm x 10 cm)
- 2: Prep Sponge Swab


2: Gauze Pad: 2" x 2" (5 cm x 5 cm)

5: Gauze Pad: 4" x 4" (10 cm x 10 cm)

1: Suture: 2-0 Silk with Curved Needle

1: Tray: Prep

Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death.
Not made with natural rubber latex.
Sterilized using ethylene oxide unless otherwise indicated in the contents list.
Fluid path components are non-pyrogenic.

 Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (16 Ga.)	0.4	2947
Medial (18 Ga.)	0.4	1552
Proximal (18 Ga.)	0.4	1702

* Priming volumes are approximate and are done without accessories.
† Flow rate values are approximate and are determined using deionized water at 100 cm head height.



Arrow International LLC
Subsidiary of Teleflex Incorporated
3015 Carrington Mill Blvd.
Morrisville, NC 27560 USA

Teleflex®
www.teleflex.com

LBL053766 R01 (2023-06)



REF MTO-12703-J

7 Fr. 3 Lumen 16 cm catheter length .032 inch dia. spring-wire guide

Three-Lumen CVC