



REF EU-15802-CVT (IPNXXXXXX)

8 Fr. 2 Lumen 20 cm catheter length .032 inch dia. spring-wire guide

# Two-Lumen CVC

**Contents:**

- 1: Two-Lumen Catheter: 8 Fr. (2.8 mm OD) x 20 cm
- 1: Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 23-5/8" (60 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer with ECG Mark
- 1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW
- 1: Syringe: 5 mL Luer-Slip
- 1: Tissue Dilator: 9 Fr. (3.0 mm) x 10.2 cm
- 1: ECG Extension Cable
- 2: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: Scalpel: #11

**Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. [www.teleflex.com/IFU](http://www.teleflex.com/IFU)**  
**Not made with natural rubber latex.**  
 Fluid path components are non-pyrogenic.

All components are CE 2797 unless otherwise noted.



CAS # 7440-48-4

en be bg hr cs da nl et fi fr de el hu is it lv lt no  
 pl pt ro ru sr sk sl es sv tr uk



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (14 Ga.)	0.74	5354
Proximal (14 Ga.)	0.81	4324

\* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

EU Authorized Representative and Importer:

Teleflex Medical  
 IDA Business and Technology Park  
 Dublin Road, Athlone  
 Co. Westmeath, Ireland  
 2797

All components are CE 2797 unless otherwise noted.

Arrow International LLC  
 Subsidiary of Teleflex Incorporated  
 3015 Carrington Mill Blvd.  
 Morrisville, NC 27560 USA

Product of xxx  
 Packaged in xxx

LotNumber

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXXXXXXXX  
 (17)YYMMDD  
 (11)YYMMDD  
 (10)LotNumber

**Teleflex®**  
[www.teleflex.com](http://www.teleflex.com)

LBL055094 R02 (2024-08)



8 Fr. 2 Lumen 20 cm catheter length .032 inch dia. spring-wire guide

REF EU-15802-CVT

# Two-Lumen CVC

