



REF EU-14955-CVT (IPNXXXXXX)

9.5 Fr. 5 Lumen 30 cm catheter length .032 inch dia. spring-wire guide

Five-Lumen CVC

Contents:

- 1: Five-Lumen Catheter: 9.5 Fr. (3.3 mm OD) x 30 cm
- 1: Spring-Wire Guide, Marked: .032" (0.81 mm) dia. x 26-3/4" (68 cm) (Straight Soft Tip on One End - "J" Tip on Other) with Arrow Advancer with ECG Mark
- 1: Introducer Needle: Echogenic 18 Ga. x 2-1/2" (6.35 cm) XTW
- 1: Syringe: 5 mL Luer-Slip
- 1: Tissue Dilator: 10 Fr. (3.5 mm) x 10.2 cm
- 1: ECG Extension Cable
- 5: Dust Cap: Non-Vented
- 1: SecondSite™ Adjustable Hub: Fastener
- 1: SecondSite™ Adjustable Hub: Catheter Clamp
- 1: Scalpel: #11


Warning: Read all package insert warnings, precautions, and instructions prior to use. Failure to do so may result in severe patient injury or death. www.teleflex.com/IFU
Not made with natural rubber latex.
 Fluid path components are non-pyrogenic.

All components are CE 2797 unless otherwise noted.



CAS # 7440-48-4

en be bg hr cs da nl et fi fr de el hu is it lv lt no
 pl pt ro ru sr sk sl es sv tr uk



Lumen	Priming Volume* (mL)	Gravity Flow Rate† (mL/hr)
Distal (16 Ga.)	0.58	2689
Medial 1 (14 Ga.)	0.69	4100
Medial 2 (18 Ga.)	0.43	863
Medial 3 (20 Ga.)	0.43	640
Proximal (18 Ga.)	0.48	921

* Priming volumes are approximate and are done without accessories.

† Flow rate values are approximate and are determined using deionized water at 100 cm head height.

EU Authorized Representative and Importer:

Teleflex Medical
 IDA Business and
 Technology Park
 Dublin Road, Athlone
 Co. Westmeath, Ireland

All components are CE 2797 unless otherwise noted.

Arrow International LLC
 Subsidiary of Teleflex Incorporated
 3015 Carrington Mill Blvd.
 Morrisville, NC 27560 USA

Product of xxx
 Packaged in xxx

LOT Lot Number

YYYY-MM-DD

YYYY-MM-DD



(01)XXXXXXXXXXXXXXXXXX
 (17)YYMMDD
 (11)YYMMDD
 (10)Lot Number

Teleflex®
www.teleflex.com

LBL055301 R02 (2025-03)



9.5 Fr. 5 Lumen 30 cm catheter length .032 inch dia. spring-wire guide

REF EU-14955-CVT

Five-Lumen CVC

